A-A-54001
10 October 1989

COMMERCIAL ITEM DESCRIPTION

SCREW, BONE, MAXILLOFACIAL SURGERY

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a maxillofacial bone screw.

Salient characteristics:

Material. The bone screw shall be fabricated from titanium 6 AL - 4V alloy in accordance with ASTM Designation F136.

Style and design. The bone screw shall be designed for use with mesh mandibular implants in oral and maxillofacial surgical implant procedures using autogenous marrow and cancellous bone discontinuity defects of the mandible (see Notes). The style and design of the bone screw shall be in accordance with Defense Personnel Support Center Drawing No. 22408.

Sterilization. The bone screw shall be capable of undergoing steam autoclaving at 270 ± 5°F and 27 ± 1 pounds per square inch and ethylene oxide sterilization without showing any signs of corrosion.

Surface preparation. Description of surface characteristics and methods of surface preparation shall be in accordance with ASTM F86.

Finish. The bone screw shall have a uniform finish. There shall be no missing thread areas or gouges.

Workmanship. The bone screw be free from defects which detract from its appearance or impair its serviceability.

Unit. Package PG). One package containing thirty screw bones, as specified, constitutes one unit.

Unit package. The screws shall be packaged in a container designed to prevent damage during shipment or storage.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination such as visual or auditory investigation without the use of special laboratory appliances or procedures and testing determination by technical means of physical and chemical properties of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance or such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component material, the lot number, and the lot size.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, at inspection level II, except dimensions shall be sampled for examination at inspection level of S-1. The unit of product for examination purposes shall be one bone screw. Examination for defects shall include but shall not be limited to those listed below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Defects

Style and design not as specified.
Screws not clearly cut.
Threads not clear, free of foreign matter.
Threads bent, deformed or defaced.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, at inspection level S-1. The unit of product for test purposes shall be one bone screw, except for chemical analysis, a sufficient quantity of raw material from each lot shall be sampled.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6515-01-163-8081	Screw, Bone, Maxillofacial Surgery, Titanium

NOTE 2: The following is a list of the maxillofacial implant pieces and their corresponding NSNs:

6515-01-163-3762	Left mandibular body
6515-01-163-3763	Left mandibular angle
6515-01-163-7935	Flat sheet
6515-01-163-8078	Right mandibular angle
6515-01-163-8079	Mandibular symphysis
6515-01-163-8082	Bone Screw Screwdriver
6515-01-164-2678	Right mandibular body

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6815-4373
LOCATION: NCR DISP 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

(FOLD BACK ALONG THIS LINE)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54001	SCREW, BONE MAXILLOFACIAL, SURGERY

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION (Mark one
	☐ VENDOR
b ADDRESS (Street, City, State, ZIP Code)	☐ USER
	☐ MANUFACTURER
	☐ OTHER (Specify)

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI – Optional)	b WORK TELEPHONE NUMBER (Inc Code – Optional
c MAILING ADDRESS (Street, City, State, ZIP Code – Optional)	8 DATE OF SUBMISSION (YYMMDD

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-28T09:27Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

Fold along this line

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL

FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54001	SCREW, BONE, MAXILLOFACIAL, SURGERY

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.