A-A-54008
27 September 1989

COMMERCIAL ITEM DESCRIPTION

LENS, DIPLOPIA TEST, Red

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers red diplopia test lens used in the diagnosis of diplopia, commonly known as double vision.

Salient characteristics.

The diplopia test lens shall consist essentially of a handle with a round or oval red test lens attached to one end.

The lens material shall be fabricated from a red transparent thermoplastic synthetic resin (See Note 2) having the following percent light transmissions as shown in Table I.

Wavelength (nanometers)	Transmission range (percent)
380	6.2 - 13.4
400	1.5 - 4.8
420	0
440	0
460	0
480	0
500	0
520	0
540	0
560	0
580	0
600	2.4 - 6.9
620	25.0 - 36.3
640	62.2 - 69.9
660	83.2 - 85.7
680	89.9 - 90.5
700	91.6 - 91.8

TABLE I. - LENS TRANSMISSION

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

The entire length of the lens and handle shall be 5 inches (12.7 cm) nonimal.

The lens shall be 1-1/2 inches (3.81 cm) nominal in diameter and 1/8 inches (0.32 cm) thick.

The handle shall be 3/4 inches (1.91 cm) wide and not less than 1/8 inches (0.32 cm) thick.

Workmanship. The red diplopia test lens shall be free from scratches warpage, pimples, cloudiness, optical distortion and other defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One red diplopia test lens, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6540-00-346-2800

Note 2: The Rohm and Haas Plexiglass No. 2423 is one of the accepted material.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS
FDA-MPQAS

Project No. 6540-1222

Location: CID/LENS/S3C/17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:27Z
Check the source to verify that this is the current version before use.