A-A-54012
14 February 1990

COMMERCIAL ITEM DESCRIPTION

CARVERS, DENTAL (PKT)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers dental carvers suitable for use in spreading and carving dental wax.

Salient Characteristics:

Classification. The carvers shall be of the following numbers, as specified below.

PKT, No. 1

PKT, No. 2

PKT, No. 3

PKT, No. 4

PKT, No. 5

Materials.

Operative ends. The operative ends of one-piece carvers shall be fabricated of Class 4, type 440A stainless steel. Three-piece carvers shall have operative ends fabricated of class 4, type 440A, 440B or 440C stainless steel. Stainless steel shall conform to ASTM Designation F899 for composition only.

Handles. Handles of one-piece carvers shall be fabricated of Class 4, type 440A stainless steel. Alternate handles No. 1 and No. 2 shall be fabricated of Class 3, type 303 stainless steel. Stainless steel shall conform to ASTM Designation F899 for composition. Alternate handle No. 3 shall be fabricated of brass. Brass shall conform to copper alloy number C26000 as specified in ASTM Designation B135 for composition. The brass handle shall be chromium over nickel plated.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Hardness. Operative ends shall have a Rockwell hardness of C52 to C56 (type 440A), C54 to C58 (type 440B) and C56 to C60 (type 440C) after proper stress relieving. Handles and shanks shall be properly tempered and shall have a hardness suitable for the purpose intended. Hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers).

Design and dimensions. Carvers shall be of Peter K. Thomas design, shall be double-ended and shall be suitable for use, with or without heat, in spreading and carving dental wax during the fabrication of dental cast restorations. The design and dimensions of the carvers shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24423 for the applicable style numbers. The carvers shall be furnished with the handle and shank design of either the one-piece instrument handle or alternate handle No. 1, No. 2 or No. 3. The carvers shall be furnished in the applicable style, as specified.

Construction. Carvers shall be double-ended and of one-piece or three-piece construction (one operative end at each end of the handle). The handle shall be of solid or hollow construction with knurled or serrated portions as specified on the drawing for the respective handle. The operative ends of three-piece carvers shall be permanently secured in the handle. The interface of the shank and handle shall be smoothly blended with no crevices evident.

Finish. The entire instrument, except for the serrated or knurled portion of the handle, shall have a mirror or bright polished finish.

Corrosion resistance. The copper sulfate test requirements of ASTM Designation F1089 shall apply to the operative ends and all stainless steel handles. The boil test requirements of ASTM Designation F1089 shall apply to the entire instrument.

Identification marking. The carvers shall be marked "PKT" with the respective carver number and name of manufacturer. The word "stainless" shall be marked on the handle. As an alternative to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer. Carvers of foreign manufacture shall be marked with the country of origin. Markings shall be on the smooth surface of the handle.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Workmanship. The operative ends shall be free from coarse grind marks, cracks, burrs, pits or feathers. The serrated or knurled portion of the handle shall be free of barbs or slivers. The entire instrument shall be free of all other defects which detract from its appearance or impair its intended use.

Unit. Each (EA). One carver of the number specified, constitutes one unit.

Unit package. Each carver shall be wrapped in neutral tissue or other suitable anti-corrosive material and in a close fitting container to prevent damage to the operative ends.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an inspection level of II. The unit of product for examination purposes shall be one carver as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Examination. Examination shall be as specified herein. Examination shall include, but shall not be limited to the following:

Defects

Style and design not as specified for the applicable style of instrument.
Finish of instrument not as specified.
Operative end not free from coarse grind marks, cracks, burrs, pits, feathers, and other imperfections.
Carvers not free from dirt, oil polishing compound, and foreign matter.
Identification marking not present, not properly located, illegible or incomplete.
Handle serrations missing, not uniform, or not complete.

Sampling for tests. Sampling for tests shall be conducted in accordance woth MIL-STD-105 and as specified below. The unit of product for test purposes shall be one carver.

Test Criteria

Characteristic	Inspection Level
Chemical analysis	*
Hardness	S-2
Dimensions	S-2
Copper Sulfate test	S-2
Boil test	S-2

* A sufficient quantity of each lot of metal used in the construction of the carvers may be selected for composition analysis.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Chemical analysis. In the absence of certificates of quality, chemical analysis of the individual components of the carver shall be determined in accordance with method 111 of FED-STD-151.

Dimensions. Dimensions shall be measured using an optical comparator, shadowgraph, or other suitable device capable of measuring the carver dimensions with the accuracy specified in the drawing.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:28Z
Check the source to verify that this is the currentversion before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine of the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-01-056-0876	CARVER, DENTAL, PKT, No. 1: Double-ended
6520-01-056-0877	CARVER, DENTAL, PKT, No. 2: Double-ended
6520-01-056-0878	CARVER, DENTAL, PKT, No. 3: Double-ended
6520-01-056-0879	CARVER, DENTAL, PKT, No. 4: Double-ended
6520-01-056-0880	CARVER, DENTAL, PKT, No. 5: Double-ended

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No.6520-2291
Location- NCR Disk 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In the continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54012	CARVERS, DENTAL (PKT)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.