A-A-54030
16 November 1989

COMMERCIAL ITEM DESCRIPTION

BRUSH, BRISTLE, DENTAL HANDPIECE (Mandrel Mounted)

The General Services Administration has authorized the use of this Commerical Item Description as a replacement for Federal Specification GG-B-735B which is cancelled.

Salient characteristics:

This commercial item description covers mandrel mounted bristle brushes suitable for polishing plastic dentures. Brushes shall be of the following types as specified below:

Type I - Stiff bristle

Type II - Soft bristle

Material:

Bristles. Type I shall be top-grade, stiff, silver gray or white natural bristles, or silver gray, horsehair. Type II shall be top-grade soft, white or gray goat's hair.

Metal components. Metal components, including mandrels, shall be fabricated of brass; or steel, nickel or chromium plated.

Style, design and dimensions. Brushes shall be of the style, design and dimensions shown in Figure 1. The bristles shall be held securely in place by a washer assembly. The washer assembly shall consist of two halves. The two halves shall be fitted together in such a manner that when the brush is used in a high-speed straight laboratory handpiece, having a no load speed of 34,000 to 40,000 RPM, there shall be no significant bristle loss when tested as specified herein. The assembly shall be mounted true on the mandrel, and permanently affixed, such that the eccentricity of the brush shall not be effected at high speeds, as a result of loosening of the washer assembly on the mandrel, even after repeated uses. The mandrel shall be suitable for use with high-speed straight laboratory dental handpieces. The brushes shall function satisfactorily when used in dental operative procedures in cleaning and polishing dental restorations and prosthetic appliances. All edges of the washer assembly shall be rounded and smooth to prevent cutting of the bristle as a result of pressure exerted on the brush during dental procedures.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Corrosion resistance. The metal components shall show no more than 2 percent corrosion of the surface area when subjected to the boil test as specified in ASTM Designation F1089.

Finish. The exposed metal components of the brush shall have a bright smooth finish.

Identification marking. Each brush shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

Workmanship. Brushes shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing twelve brushes of the type specified, constitutes one unit.

Unit package. Twelve brushes of each type shall be packaged in a cardboard box. A separate piece of cardboard shall be perforated with holes to accomodate and secure the mandrels of the brushes, inside the cardboard box.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination and test characteristics. The unit of product for inspection, examination and test purposes shall be one brush for each type. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, and sampling for tests (when applicable shall be conducted at inspection level S-2).

Examination. Examination shall include, but shall not be limited to the following:

Defects

Type not as specified.
Style and design of brush not as shown in Figure 1.
Bristles not securely fastened in the washer assembly.
Washer assembly not permanently mounted on the mandrel.
Finish not as specified.
Color of bristles not as specified.
Dimensions not as specified.

Tests.

Chemical analysis. In the absence of certificates of quality, the chemical analysis of the instrument shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of a conflict, method 111 shall govern.

Bristle test. Brushes shall be tested using a wet pumice abrasive on a plastic denture. The brushes shall be run for a minimum of five 2 minute intervals, where at the end of each interval the brushes shall be rinsed off with clean water and inspected for bristle loss.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and of regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification	Type
6520-01-056-7375	BRUSH, BRISTLE, DENTAL HANDPIECE, Mandrel Mounted, 3/4 inch, Soft bristle, 12s	II
6520-01-056-7376	BRUSH, BRISTLE, DENTAL HANDPIECE, Mandrel Mounted, 3/4 inch, Stiff bristle, 12s	I

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB


AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS

PROJECT NO. 6520-2314
LOCATION: NCR - DISK 8

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.



NOTES:

DESIGN OF BRISTLE HOLDING COMPONENTS IS FOR ILLUSTRATION PURPOSES ONLY.

ALL DIMENSIONS ARE IN INCHES.

FIGURE 1.

MANDREL MOUNTED BRISTLE BRUSH

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.