A-A-30129A
February 28, 1992
SUPERSEDING
A-A-30129
November 5, 1987

COMMERCIAL ITEM DESCRIPTION

BANDAGE, GAUZE, CALAMINE-GELATIN-GLYCERIN-ZINC OXIDE IMPREGNATED

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes a calamine-gelatin-glycerin-zinc oxide impregnated gauze bandage.

Salient characteristics. Bandage, Gauze, Calamine-Gelatin-Glycerin-Zinc Oxide Impregnated, nominal 4 inches wide by 10 yards long (10.16cm by 9.14m), shall have a nominal weight of 30 grams, rolled, white, gauze 100% cotton, may have woven edge or looped edge and shall be free of lint and excessive fraying (nonravelling), warp 28 to 30 ends per inch, filling 9 to 11 double picks per inch; each bandage shall be impregnated with not less than 175 grams and not more than 230 grams of calamine-gelatin-glycerin-zinc oxide containing paste, per roll. Zinc oxide content shall be not less than 3.5% and not more than 10.1% (w/w) of the calamine-gelatin-glycerin-zinc oxide impregnated gauze bandage.

Microbial limit - bandage shall contain no vegetative cells and shall contain not more than 50 aerobic spore-forming organisms when tested using U.S.P. microbial limit test.

The finished product shall contain suitable preservaties.

Moisture - the bandage shall have sufficient moisture for smooth application and shall adhere to the area being wrapped and to itself when allowed to dry.

There must be less than 100 colony forming units per gram; there must be no pathogens and no coliforms.

Workmanship. The material and its containers shall be free from defects which detract from their appearance or impair their serviceability. All ingredients entering into the finished product shall be of a pharmaceutical grade.

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One each, as specified, constitutes one unit. The unit for the VA procurement shall appear in the solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 and 2.5 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one bandage gauze, calamine- gelatin glycerin - zinc oxide impregnated as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	NOMENCLATURE
6510-00-450-5605	BANDAGE, GAUZE, CALAMINE-GELATIN-GLYCERIN-ZINC OXIDE IMPREGNATED, 4 inches by 10 yards

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Copies of Federal Test Method Standard 191 may be obtained from your local General Services Administration

Copies of United States Pharmacopedia (USP) regulations may be obtained from USPC, Inc., 12601 Twinbrook Parkway Rockville, MD 20852 USA.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.