A-A-51360A
30 August 1991
SUPERSEDING
A-A-51360
26 March 1986

COMMERCIAL ITEM DESCRIPTION

FLASK, VOLUMETRIC(Borosilicate Glass)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers precision grade single capacity line borosilicate glass volumetric stoppered flasks.

Salient Characteristics:

Shall be precision grade stoppered volumetric flasks for use in laboratory procedures.

Shall be fabricated of borosilicate glass conforming to the requirements in ASTM Designation E438.

Shall be in accordance with all applicable requirements in ASTM Designation E288 for Class A flasks and the milliliter (ml) capacities that are listed.

Shall be supplied with borosilicate ground glass pennyhead or barrelhead stoppers conforming to taper specifications outlined in ASTM Designation E675, and also supplied in stopper sizes as required by ASTM Designation E288.

Shall be tested for volumetric accuracy to comply with the requirements of the College of American Pathologists (see Quality Assurance Provisions).

Workmanship. The stoppered flasks shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One flask of the capacity specified, constitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Volumetric accuracy testing. To meet requirements for certification of the College of American Pathologists, the flasks shall be tested in accordance with ASTM Designation E542 to determine conformance to capacity tolerance requirements contained in ASTM Designation E288. A Certificate of Quality may be furnished in lieu of actual performance of individual testing, provided that lot/sample identities can be maintained. The Certificate of Quality shall include the lot number, lot size, sample size and test method.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

ASTM

ASTM E288 Standard Specification for Volumetric Flasks
ASTM E438 Standard Specification for Glass in Laboratory Apparatus
ASTM E542 Practice for Calibration of Volumetric Ware
ASTM E675 Standard Specification for Interchangeable Taper-Ground Stopclocks and Stoppers

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

For information only.

The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6640-00-368-6766	FLASK, VOLUMETRIC, Borosilicate Glass, 10 ml
6640-00-425-4000	FLASK, VOLUMETRIC, Borosilicate Glass, 25 ml
6640-00-425-5000	FLASK, VOLUMETRIC, Borosilicate Glass, 50 ml
6640-00-425-6000	FLASK, VOLUMETRIC, Borosilicate Glass, 100 ml
6640-00-425-7000	FLASK, VOLUMETRIC, Borosilicate Glass, 250 ml
6640-00-425-8000	FLASK, VOLUMETRIC, Borosilicate Glass, 500 ml
6640-00-425-9000	FLASK, VOLUMETRIC, Borosilicate Glass, 1000 ml
6640-00-426-0000	FLASK, VOLUMETRIC, Borosilicate Glass, 2000 ml

Ordering data. (Capacity required, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6640-1730

Location: ENABLE/FLASKCID/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51360A	30 August 1991

3. DOCUMENT TITLE

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215) 737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.