A-A-51885
23 MARCH 1987

COMMERCIAL ITEM DESCRIPTION

DRAIN, SURGICAL (Rubber, Mushroom Tip, Pezzer)

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification ZZ-C-00101b dated 1 August 1963.

This commercial item description covers rubber surgical drains of the Pezzer design suitable for use in major surgical procedures.

Salient characteristics:

Style and design. The proximal end of the drain shall be flared to accommodate a tubing adapter or irrigation syringe. The extreme distal tip shall be reinforced with a woven material. The distal end shall be a mushroom head of the Pezzer design with two to four holes. The holes shall have a cross sectional area equal to or greater than that of the lumen of the drain. The drain shall have dimensions shown in Table I for the size specified. The shaft outside diameter shall be the label French size (+0.067 cm/-0.033 cm).

Table I. Drain characteristics.

French size	Inside Diameter (cm) Minimum	Mushroom (cm) Minimum	Head Diameter (cm) Maximum	Break Strength (pounds) Minimum	Ultimate Elongation (percent) Minimum
16	0.250	1.3	1.7	28	750
18	0.285	1.4	1.7	36	750
20	0.355	1.5	2.0	44	750
24	0.463	1.9	2.2	60	725
26	0.483	1.9	2.6	68	725
28	0.516	1.9	2.6	74	725
30	0.533	1.9	2.6	82	725
32	0.600	1.9	3.2	90	725

Markings. The drain shall bear the manufacturer's name or registered trademark and the French size on the proximal end.

Material. The drain shall be manufactured from natural latex rubber.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the current version before use.

Biocompatibility testing. Material used in the manufacture of the drains shall be tested in accordance with the Intramusular Implantation and Intracutaneous tests of the USP.

Workmanship. The exterior and lumen of the drain shall be free from cracks, blisters, embedded particles, and foreign matter. There shall be no defects which will detract from the appearance or impair the serviceability of the drain.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). Each drain of the size specified shall be packaged clean, uncoiled in a sealed peel-open type container.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Certificates of Biocompatibility. A certificate of compliance, stipulating that offered items meet the biocompatibility test, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years, and that there have been no changes to material formulation and method of processing since actual testing was performed. The certificate shall include the date of actual testing, the test method and results, the name and title of the person evaluating results, the current contract number and NSN, a statement to the effect that material processing and formulation have not changed since actual testing was performed, and the signature and title of the person making the certification.

Sampling for physical property testing. Sampling shall be conducted in accordance with MIL-STD-105, inspection level S-1, AQL 1.0. The unit of product for test purposes shall be one drain.

Physical properties. The drain shall be tested for breaking strength and ultimate elongation using paragraphs 1 thru 4 of Federal Test Methods 4111 and 4121 respectively of Federal Test Method Standard Number 601, except that two specimens shall be cut from each drain. One specimen shall be tested aged and one specimen unaged. The test value from each specimen shall constitute one test value. Breaking strength shall be the value of the breaking force, as defined in Federal Test Method 4111. The breaking strength and ultimate elongation shall be as shown in Table I for the size specified. In addition, the drain shall be subjected to accelerated aging in accordance with Federal Test Method 7121 of Standard Number 601. After aging, the drain shall not become brittle or tacky, and the tests for breaking strength and ultimate elongation shall be repeated. The values for the aged drains shall not be less than 75 percent of the values obtained for the unaged drains.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-299-8751	DRAIN, SURGICAL, Rubber, Mushroom Tip, Pezzer, 16 Fr
6515-00-299-8752	DRAIN, SURGICAL, Rubber, Mushroom Tip, Pezzer, 18 Fr
6515-00-317-8550	DRAIN, SURGICAL, Rubber, Mushroom Tip, Pezzer, 20 Fr
6515-00-317-8570	DRAIN, SURGICAL, Rubber, Mushroom Tip, Pezzer, 24 Fr
6515-00-317-8580	DRAIN, SURGICAL, Rubber, Mushroom Tip, Pezzer, 26 Fr
6515-00-317-8590	DRAIN, SURGICAL, Rubber, Mushroom Tip, Pezzer, 28 Fr
6515-00-317-8610	DRAIN, SURGICAL, Rubber, Mushroom Tip, Pezzer, 30 Fr
6515-00-317-8630	DRAIN, SURGICAL, Rubber, Mushroom Tip, Pezzer, 32 Fr

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PBS
FDA-MPQAS

PGC 05044

Project No. 6515-3551

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4964 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51885	DRAIN, SURGICAL (RUBBER, MUSLIN TYP, PEZZ)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.