A-A-51906A
31 August 1990
SUPERSEDING
A-A-0051906(DM)
7 APRIL 1987

COMMERCIAL ITEM DESCRIPTION

TUBE, INHALER

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a oxygen connecting tube used in respiratory procedures.

Salient characteristics:

General. The oxygen tubing shall be compatible for use with rebreathing bags, hand held nebulizers and other oxygen administration devices.

Dimensions. The oxygen tubing shall be 72 ± 3 inches long. The inside diameter shall be 9/64 ± 1/64 a inch.

Materials. The tubing shall be nonconductive, and transparent with a green tint. It shall be manufactured from a flexible PVC plastic, and the female connectors shall be manufactured from a rigid plastic.

Style and Design. The outside tubing surface may either be smooth or corrugated. The tubing ends shall have two permanently attached molded, resilient, tapered female connectors. The connectors shall make a secure and leakproof connection with tubing connectors of outside diameter 0.9 to 1.1 cm.

Vacuum test. The tubing shall not collapse when subjected to an internal vacuum of minimum 40 cm of mercury for 30 seconds. Tubing shall be tested for vacuum resistance. Acceptance number shall be zero (0).

Female connector test. Attaching the tubing to a standard male adapter of outside diameter graduated from 0.9 to 1.1 cm, a five pound tensile force shall be applied to the tubing for 30 seconds. The tubing shall not separate from the connector while the force is being applied.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Workmanship. The tubing shall be clean and free from foreign matter, cracks and kinks. The tubing shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty individually packaged oxygen tubings, as specified, constitutes one unit. The individual package shall be a sealed, peel open type container capable of maintaining cleanliness of the item and of preventing kinking of the tubing.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. The tubing shall be examined to determine compliance with specification requirements. Sampling shall be in accordance with MIL-STD-105, sampling level II, AQL 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the currentversion before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one coil of tubing.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-562-8308	TUBE, INHALER, OXYGEN, PLASTIC, 72 Inch Long, 50's

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

A-A-51906A

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4559

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.