A-A-51908A
14 February 1989
SUPERSEDING
A-A-0051908(DM)
3 November 1987

COMMERCIAL ITEM DESCRIPTION

FORCEPS, ENDOSCOPIC, (Large, Forward Grasping, Jackson)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers forward grasping endoscopic forceps of the Jackson design, used in endoscopes 5 mm or larger.

Salient characteristics:

General. The forceps shall be a large size cannulated endoscopic forceps of the Jackson design, used in endoscopic procedures. The box lock instrument shall have jaws which are shaped and set to provide a secure grip on the object being removed. The forceps shall be capable of being used through endoscopes 5 mm or larger.

Material. The forceps shall be fabricated from Class 4 martensitic corrosion-resistant steel in accordance with ASTM Designation F899.

Hardness. The forceps shall have a Rockwell Hardness C scale (HRC) of 48 to 53 in accordance with ASTM Designations E18 or E92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Designation E140.

Style, design and dimensions. The forceps shall be 50 or 60 cm long and have rounded jaws whose faces are serrated and grooved. The forceps shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24452.

Passivation. The forceps shall be passivated.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Finish. The forceps shall be mirror, satin, or matte. The matte finish shall exhibit a uniform, non-directional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Corrosion-resistance. The forceps shall show no signs of corrosion or plating of copper when tested for corrosion-resistance in accordance with ASTM Designation F1089.

Performance. When fully closed, the jaws shall interdigitate. Action of the forceps shall be easy, without drag or binding. Performance shall be in accordance with applicable portions of ASTM Designation F1026.

Identification marking. The forceps shall bear the contractor's name or registered trademark and the word "stainless". Forceps of foreign manufacture shall bear the name of the country of manufacture.

Workmanship. The forceps shall be free from burrs, cracks, sharp edges, scale, corrosion or other defects which detract from its appearance or impair its serviceability.

Product Quality Level

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One forceps, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Sampling for inspection shall be level S-2. Dimensional testing shall be at a 2.5% AQL, hardness testing at a 1.5% AQL, and corrosion-resistance testing at a 1.0% AQL. The unit of product for test purposes shall be one forceps.

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-117-3878	FORCEPS, ENDOSCOPIC, Large, Forward Grasping, Jackson, 50 cm Long
6515-01-139-8379	FORCEPS, ENDOSCOPIC, Large, Forward Grasping, Jackson, 60 cm Long

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4548

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the currentversion before use.