A-A-51909A
14 February 1989
SUPERSEDING
A-A-0051909(DM)
15 September 1987

COMMERCIAL ITEM DESCRIPTION

FORCEPS, ENDOSCOPIC (Alligator, Grasping, Jackson)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers alligator-style grasping forceps of the Jackson design.

Salient characteristics:

General. The forceps shall be a Jackson design, alligator style grasping forceps, used in endoscopic procedures. The forceps shall be designed to pass through all laryngoscopes and bronchoscopes 6 mm or larger.

Material. The forceps shall be fabricated from Class 4 martensitic stainless steel in accordance with ASTM Designation F899.

Hardness. The forceps shall have a Rockwell Hardness C Scale (HRC) of 48 to 52 in accordance with ASTM Designations E18 or E92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Designation E140.

Style, design and dimensions. The forceps shall be of rigid construction, 50 or 60 cm long with serrated and grooved jaws 4 mm wide and 22 mm long. The forceps shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24459.

Passivation. The finished forceps shall be passivated.

Identification markings and labeling. The forceps shall bear the contractor's name or registered trademark and the word "stainless". Forceps of foreign manufacture shall bear the name of the country of manufacture.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Corrosion-resistance. The forceps shall show no signs of corrosion or plating of copper when tested for corrosion resistance in accordance with ASTM Designation F1089.

Finish. The forceps shall be mirror, satin, or matte. The matte finish shall exhibit a uniform, non-directional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Performance. The action of the forceps shall be easy, without drag or binding. The jaw teeth shall interdigitate in the fully closed position. The shanks shall be constructed so they will not separate during use.

Workmanship. The forceps shall be free from burrs, cracks, sharp edges, scale, corrosion, or other defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One forceps, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Sampling for inspection shall be level S-2. Dimensional testing shall be at 2.5% AQL, hardness testing at a 1.5% AQL, and corrosion resistance testing at a 1.0% AQL. The unit of product for test purposes shall be one forceps.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-117-3877	FORCEPS, ENDOSCOPIC, Alligator, Grasping, Jackson, 50 cm Long
6515-01-139-8376	FORCEPS, ENDOSCOPIC, Alligator, Grasping, Jackson, 60 cm Long

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4547

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.