A-A-0051915(DM)
10 April 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

TESTER, DEFIBRILLATOR ENERGY

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers a defibrillator energy tester suitable for use in verifying or calibrating the output energy from all DC defibrillators.

Salient characteristics:

The defibrillator energy tester shall operate from a nominal 115 VAC, 60 Hz, or battery power. If it is battery powered, it shall have a battery life of at least one year. It shall be furnished with a suitable hard protective carrying case. The tester shall measure trapezoidal or LOWN wave forms of 2 to 40 mS duration. It shall have an operating temperature range of 15 to 40°C, an energy measurement range of 0 to 800 Joules, ±4 percent and a test load of 50 ohms, ± 1 percent. The tester shall measure waveforms of up to 5 kV amplitude. It shall have a real time oscilloscope output with a BNC or banana type connector, and an attenuation of 1000:1. The tester shall weigh less than 9 pounds, and measure no more than 9" x 10" x 15". It shall have a digital or analog energy scale and an energy readout reset.

Workmanship. The defibrillator energy tester shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One defibrillator energy tester, with carrying case, as specified, constitutes one unit.

AMSC N/A FSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for testing and examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one defibrillator energy tester.

TABLE I. Sampling for testing and examination.

	Inspection level	AQL (defects per hundred units)
Major defects	III	0.4
Minor defects	III	0.4

Tests. All tests shall be performed on a first article sample in accordance with table II.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the currentversion before use.

*Examination shall not be limited to the classified possible defects listed herein.

Table II. Tests.

1) Safety. If the unit is line powered, electrical leakage current shall comply with those limits specified in ANSI/AAMI SCL 12/78. Tests shall be in accordance with above standard. Equipment shall be a leakage current tester/safety analyzer with capability to perform tests described in above standard.
2) Performance. The manufacturer shall demonstrate to the Government via mutually acceptable test methods, compliance with the following specifications and performance criteria: Waveforms - trapezoidal or LOWN, 2 to 40 mS duration. Energy measurement range -0 to 800 joules, ±4 percent. Test load - 50 ohms, ±1 percent. Waveform amplitude - 5kV.
*Examination. Each defibrillator energy tester shall be examined to determine compliance with the classification of defects in table III.

TABLE III. Categories and Defects *

Categories	Defects
Major	
101	Controls(s) non-functional.
102	Indicator non-functional.
103	Output(s) non-functional.
Minor	
201	Identification marking(s) not complete, permanent, and legible.
202	Finish not free of chips, cracks, scratches, and other appearance defects.
203	Any component part not secured in place.

*Examination shall not be limited to the classified possible defects listed herein.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6625-00-433-9063

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6625-D888

PGC H1085

Location: CID/TESTER/H17-22

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.