COMMERCIAL ITEM DESCRIPTION

SLIDE MICROSCOPE (Plain, Frosted End)

This commercial item description covers a precleaned, straight edge, frosted end 75 X 25 millimeter(mm) microscope slide that is 1 mm thick, for use in laboratory procedures involving blood smear work or the mounting of histological and cytological preparations.

Salient characteristics:

Shall be a 1 mm thick 75 X 25 mm glass microscope slide with a frosted end (i.e. with marking sector), for use in laboratory procedures.

Shall have dimensions of 75 mm length X 25 mm width, with a $\pm$ 0.8 mm tolerance for each dimension. The allowable thickness range shall be 0.8 to 1.1 mm.

Shall be fabricated from clear transparent glass, free from imbedded foreign material, bubbles, seeds, blisters, striae and internal cloudiness.

Shall be supplied precleaned, free of residue, and ready for use. The faces of the microscope slide shall be such that, after being dipped into a beaker of distilled water, both surfaces immediately after removal from the water exhibit a uniform dispersion of liquid and absence of smears.

Shall have all four edges flat and perpendicular to the microscope slide faces without any beveling. The edges of the short dimension shall be free of chips, and shall be finished to ensure smoothness and uniformity for use in blood smear work. The edges of the long dimension shall be straight and shall be free of roughness or sharpness to facilitate handling. The corners of the microscope slide shall be square.

Shall be supplied with a single or double sided frosted end to give 17 mm to 21 mm marking sectors(s) that allow legible markings to be applied with a No. 2 lead pencil. The frosting shall be sufficient to hold the No. 2 lead pencil markings through normal rubbing, cytology fixatives (water and alcohol based), and submersion in alcohol and xylene.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the current version before use.

Shall have microscope faces free of scratches which will affect its use.

Shall be suitable for mounting histological and cytological preparations. Tissue sections shall adhere to the microscope slide which has been prepared with a proper binding agent, such as gelatin or albumin, prior to the affixing of the tissues.

Expiration dating period. An expiration dating period shall be applied to the unit packaging. The expiration dating period shall be 24 months from the date of manufacture, representing the date of unit packaging.

Date of delivery. Not less than 21 months of the expiration dating period shall remain at time of delivery to the Government.

Workmanship. The microscope slide shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package with a heat sealed overwrap containing 72 slides with an inner wrap, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-00-074-4191

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1404

Location: Disk # 5

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER
A-A-51919

2. DOCUMENT TITLE
SLIDE MICROSCOPE (Plain, Frosted End)

3a. NAME OF SUBMITTING ORGANIZATION

3b. ADDRESS (Street, City, State, ZIP Code)

4. TYPE OF ORGANIZATION (Mark one)

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER (Specify):

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) - Optional

7b. WORK TELEPHONE NUMBER (Include Area Code) - Optional

7c. MAILING ADDRESS (Street, City, State, ZIP Code) - Optional

8. DATE OF SUBMISSION (YYMMDD)

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a [illegible] effort to make our standardization [illegible] better, the DoD provides this form for use [illegible] and suggestions for improvements. All users of military standardization [illegible] are invited to provide suggestions. This form may be [illegible], folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, too rigid, [illegible], loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4868 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.