A-A-51938A
10 September 1990
SUPERSEDING
A-A-51938
24 April 1987

COMMERCIAL ITEM DESCRIPTION

CASSETTE, RADIOGRAPHIC FILM

(Lightweight)

The General Services Administration has authorized the use of this commercial item description as a replacement for A-A-51249B which is cancelled

This Commercial Item Description covers a lightweight radiographic film cassette with two high speed intensifying screens

Salient Characteristics:

The radiographic film cassette shall be lightweight and suitable for use with medical x-ray film The cassette shall be furnished with two high speed intensifying screens The cassette shall be furnished for the following nominal film sizes:

Size 1	-	35 cm by 43 cm
Size 2	-	30 cm by 35 cm
Size 3	-	24 cm by 30 cm
Size 4	-	18 cm by 24 cm
Size 5	-	8 inches by 10 inches
Size 6	-	24 cm by 24 cm

The cassette shall consist of two aluminum covers mounted in an aluminum, American Iron and Steel Institute (AISI) Series 300 corrosion-resisting steel, or polyurethane frame joined together on the long dimension by a suitable hinge, or hinges. The cassette tube side (front side) shall have a scratch resistant vinyl covering. On the side opposite the hinge, there shall be a suitable fail-safe, flush mounted,

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

latching system

The cassette shall be designed to insure the proper tightness and contact of the film against the screens. One cover shall have a foam, or felt, backing to enhance this tightness. The back cover shall contain sufficient lead to preclude excess scattered radiation. The back cover of the cassette shall have a window with a latched cover in order to perform identification procedures in daylight areas.

The intensifying screens shall be in accordance with the latest revision of ANSI Standard PH1.50, Dimensions for Radiographic Film Cassette Intensifying Screens. This commercial item description shall govern in the event it deviates from the ANSI Standard.

Other specifications for the medical radiographic film cassette shall conform to ANSI Standard PH1.49, Specifications for Medical Radiographic Film Cassettes. This commercial item description shall govern in the event it deviates from the ANSI Standard.

Each cassette shall be suitable for use with the corresponding film size.

The cassette finish shall be free of chips, cracks, scratches, and other appearance defects. The cassette shall be furnished clean and free of dirt, oil, and grease. Appropriate instructions and identification stickers shall be included if available.

Workmanship The radiographic film cassette shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One radiographic film cassette of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:03Z
Check the source to verify that this is the current version before use.

suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one radiographic film cassette

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3 The unit of product for test purposes shall be one radiographic film cassette

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

Test radiographs shall be made at a focal spot to film distance of 40 plus or minus 1 inch, at 80 Kvp, with screen-type, medium speed x-ray film. The x-ray rube shall be perpendicular to the center of the cassette The nominal focal spot size of the x-ray tube shall not be more than 1.0 mm. A minimum of 2 3 mm of aluminum filtration shall be in the useful beam. The exposed x-ray film shall be developed in accordance with the

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T09:03Z
Check the source to verify that this is the currentversion before use.

manufacturer's instructions. The developed x-ray film shall be viewed with an unaided eye using a fluorescent type x-ray film illuminator.

Drop test. The selected sample shall be dropped on each corner from a height of 30 inches onto a concrete or tile floor The cassette shall be dropped so that the plane of the cassette is perpendicular to the floor and the center of gravity of the cassette is directly over the corner which strikes the floor. Except for deformation at the impacted corner area, the cassette shall show no other sign of damage. Any other kind of damage, will constitute a failure After the drop test, the sample cassette shall be subjected to and pass the flatness test, lightproof test, and film contact test.

Flatness test The test for flatness shall be conducted on the open and closed cassette. The cassette frame shall rest on a flat test surface and any height variation shall be measured by passing feeler gages between the flat surface and the frame The deviation from flatness of the front panel shall be determined by measuring the maximum distance between the front face and a straight edge placed at various angles on the front panel The damaged corners resulting from the drop test shall not be placed on the flat test surface, and shall not be measured Both frame and front panel shall be flat within 1/16 inch Any deviation will constitute a failure.

X-Ray opacity of the front panel test A non-screen type radiographic film shall be exposed in a cardboard, or plastic film holder, through the front panel with an exposure time to produce a diffuse transmission density of 1 0 plus or minus 0 2. The developed film shall be examined for uniformity of opacity and for evidence of included foreign matter. Visible shadows or non-uniformity shall be cause for rejection.

Lightproof test A high speed radiographic film, of the applicable size, shall be placed in the cassette, and all sides and edges of the cassette shall be exposed to a 100 watt incandescent lamp, operated at its rated voltage, at a distance of 3.3 feet (1 meter) from the cassette for a period of ten minutes. The developed test film shall be compared with a developed, unexposed film of the same lot A density on the test film greater than that on unexposed film shall be cause for rejection. When viewed against the illuminator, the test film shall show no density variations, except that fog within 1/8 inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

of film border or within 1/4 inch square at the film corners shall not be considered a failure.

Cassette liners test A radiographic film in the sample cassette shall be exposed with an exposure time to produce a diffuse transmission density of 1.0 plus or minus 0.2. Shadows and artifacts on the test film shall be cause for rejection of the cassette.

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented be Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTES: 1 The Intensifying Screen specified herein is equivalent to Kodak X-Omatic Regular Screen.

2 The following National Stock Numbers are covered by this document

6525-01-012-9686 (Size 1)
6525-01-012-9687 (Size 2)
6525-01-012-9688 (Size 3)
6525-01-012-9689 (Size 4)
6525-01-012-9690 (Size 5)
6525-01-111-5051 (Size 6)

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY.

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6525-1262

Location: FILMCID.13A

Source: https://assist.dla.mil -- Downloaded: 2026-04-26T09:03Z
Check the source to verify that this is the current version before use.