INCH-POUND

A-A-50909C
31 May 1994
SUPERSEDING
A-A-50909B
05 August 1991

COMMERCIAL ITEM DESCRIPTION

TUBES, DRAINAGE, SURGICAL (Penrose Type)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers Penrose type surgical drainage tubes suitable for drainage during or subsequent to surgical procedures.

Salient characteristics:

General. Except as specified, the tube shall be in accordance with American Society for Testing and Materials (ASTM) Designation D3579, Standard Specification for Rubber Surgical Drainage Tubes, Penrose Type. The tube shall be sterile, disposable and used during or after surgery.

Material. The drainage tube shall be fabricated from amber colored, x-ray opaque, latex rubber.

Style, design and dimensions. The tubes shall be of the following sizes and dimensions:

Size	Nominal ID inches	Length inches (Nom.)	Wall thickness ± 0.003 inches	Flat width ± 3/64 inch
1	3/8	12	0.011	37/64
2	5/8	12	0.011	61/64
3	7/8	12	0.011	1-21/64
4	1	18	0.011	1-1/2
5	3/8	18	0.011	37/64
6	5/8	18	0.011	61/64
7	7/8	18	0.011	1-21/64

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Sterility. The tubes shall be sterile in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.000001.

Marking. Each individual package shall contain the following information:

The measurements of the drain.
The manufacturer's name and date of manufacture or date of expiration.

Workmanship. Each tube shall be processed in such a manner as to be uniform in quality and free from any defects that will affect life, serviceability or appearance.

Unit. Package (PG). For sizes 1, 2, 3 and 4. Each drainage tube shall be individually packaged in a sealed peel-open container capable of maintaining sterility of contents unless package is damaged or opened. One package containing 200 individually packaged drainage tubes, as specified, constitutes one unit.

Each (EA). For sizes 5, 6, and 7. Each drainage tube shall be individually packaged in a sealed peel-open container capable of maintaining sterility of contents unless package is damaged or opened. One tube, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one tube as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one tube. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications and Sizes are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6515-01-385-2013	TUBE, DRAINAGE SURGICAL, PENROSE, Sterile, Disposable, 3/8 Inch ID, 12 Inches Long, 200s	1
6515-01-384-9023	TUBE, DRAINAGE SURGICAL, PENROSE, Sterile, Disposable, 5/8 Inch ID, 12 Inches Long, 200s	2
6515-01-385-1697	TUBE, DRAINAGE SURGICAL, PENROSE, Sterile, Disposable, 7/8 Inch ID, 12 Inches Long, 200s	3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6515-01-188-5316	TUBE, DRAINAGE SURGICAL, PENROSE, Sterile, Disposable, 1 Inch ID, 18 Inches Long, 200s	4
6515-00-902-1975	TUBE, DRAINAGE SURGICAL, PENROSE, Sterile, Disposable, 3/8 Inch ID, 18 Inches Long	5
6515-00-782-1544	TUBE, DRAINAGE SURGICAL, PENROSE, Sterile, Disposable, 5/8 Inch ID, 18 Inches Long	6
6515-00-782-1545	TUBE, DRAINAGE SURGICAL, PENROSE, Sterile, Disposable, 7/8 Inch ID, 18 Inches Long	7

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5707
Location: c:\nish\tubes\sg

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-50509C	31 May 1994

3. DOCUMENT TITLE
TUBES, DRAINAGE, SURGICAL (Penrose Type)

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.