INCH-POUND

A-A-54874
21 May 1993

COMMERCIAL ITEM DESCRIPTION

BANDAGE, ELASTIC, RUBBER WARP THREADS

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-B-36998B which is cancelled.

This Commercial Item Description covers elastic bandages with rubber warp threads.

Salient characteristics:

Shall be rubber elastic bandage suitable for use as a pressure bandage, support, or binder in the treatment of orthopedic and post-operative patients.

The length of the unstretched bandage shall be 57 inches minimum and the stretched length shall be a minimum average of 162 inches. As an alternate, the length of the unstretched bandage shall be 67 inches minimum and the stretched length shall be a minimum average of 180 inches.

Unstretched bandages shall be of the following widths:

SIZE	WIDTH
1	2 inches (5.08 cm)
2	3 inches (7.62 cm)
3	4 inches (10.16 cm)
4	6 inches (15.24 cm)

FSC 6510

AMSC N/A

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

The bandages shall have a minimum elasticity of 120 percent and shall be capable of returning a minimum of 110 percent of the original length. This shall be demonstrated by measuring a 10 inch length of bandage and stretching it to 22 inches for one hour, then releasing. The marked off portion of the bandage shall be remeasured. After subjection to 5 laundering cycles at 212°F and 5 drying cycles at 220°F, the bandage shall have a minimum elasticity of 120 percent and a minimum return of 75 percent. The cut ends shall be sewn or otherwise treated to prevent raveling. The edges shall be selvage. Each bandage shall be packaged with a minimum of two fastening clips made of corrosion resisting metal. If clips are sewn to the end of the bandage, they shall not loosen during laundering. Each bandage shall be one continuous piece and evenly and neatly rolled without tension. Bandages shall be flesh, beige or off-white color and shall not be greige goods, but shall be finished and shall not contain substances that are toxic or irritating to the skin. Bandages shall show good colorfastness to water.

The following styles shall be acceptable:

Style A - Bandages shall be woven having a minimum of 30 cotton yarns per inch in the warp and a minimum of 10 heat resistant rubber strands per inch in the warp covered by cotton, multi-filament polyester or multi-filament nylon suitable for the intended purpose, and a minimum of 60 cotton yarns per inch in the filling. The minimum weight per bandage shall be as follows:

SIZE	Minimum Bandage Weight (Ounce) including bandage fasteners
1	0.88
2	1.32
3	1.76
4	2.60

The minimum breaking strength shall be 20 pounds in the warp.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:39Z
Check the source to verify that this is the current version before use.

Style B - Bandages shall be knitted having a minimum of 10 rayon yarns per inch in the warp and a minimum of 10 covered heat resistant rubber strands per inch in the warp. The rubber strands shall have an inner covering of nylon and an outer covering of cotton. The filling shall contain a minimum of 54 cotton yarns per inch. Each selvage may contain one nylon warp yarn. The minimum warp breaking strength shall be 20 pounds. The minimum weight per bandage shall be as shown above.

Style C - The bandages shall be knitted having a minimum of 6 textured set polyester yarns per inch in the warp and a minimum of 6 heat resistant rubber strands per inch in the warp. The rubber strands shall be either covered with polyester or the bandage shall be constructed in such a fashion that bare rubber strands shall be covered by polyester yarns. The filling shall contain a minimum of 20 picks per inch, each pick consisting of four textured set polyester yarns.

The minimum weight per bandage shall be as follows:

SIZE	Minimum Bandage Weight (Ounce) including bandage fasteners
1	0.48
2	0.73
3	0.97
4	1.45

The minimum warp breaking strength shall be 13 pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Style D - The bandages shall be knitted having a minimum of 12 textured set polyester yarns per inch in the warp and a minimum of 12 heat resistant rubber strands per inch in the warp. The rubber strands shall be either covered with polyester or the bandages shall be constructed in such a fashion that bare rubber strands shall be covered by polyester yarns. The filling shall contain a minimum of 20 picks per inch, each pick consisting of four cotton yarns. The minimum weight per bandage shall be as follows:

SIZE	Weight (Ounce) including bandage fasteners
1	0.88
2	1.32
3	1.76
4	2.60

The minimum warp braking strength shall be 20 pounds.

In lieu of fastening clips, the fastening of the bandage shall be accomplished by the mushroom hook part of a mushroom hook and pile closure. The mushroom hook part shall be sewn across the width of one end of the bandage.

The minimum width of the mushroom hook part shall be as follows:

SIZE	Minimum Width of Hook Closure (Inch)
1	5/8
2	5/8
3	7/8
4	7/8

The mushroom hook part shall be one piece, except for size 4, the mushroom hook part shall be two pieces sewn in an abutting position across the width of the bandage. There shall be a small gap between each piece of the mushroom hook part at their abutment at the approximate center of the width of the bandage. Both before and after the five laundering and drying cycles, the mushroom hook shall firmly fasten to any part of the bandage with the bandage in the stretched condition, no matter what the degree of stretch. With the bandage wrapped around a part of the body, the mushroom hook shall be capable of being fastened and unfastened with one hand. The mushroom hook shall not loosen on the bandage during laundering.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Workmanship. Bandages shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 12 individually wrapped bandages, as specified, constitutes one unit. Each wrapped bandage shall include a minimum of two fastening clips. When the alternate bandage fastened by a mushroom hook (of a hook and pile closure) is supplied, fastening clips are not required.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination and tests. Sampling for examination and tests shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for examination purposes shall be one package of bandages.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more text characteristics. Tests shall include, but shall not be limited to, the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Length, unstretched. The bandage, without tension, shall be laid out, on a smooth horizontal surface. The length shall be measured parallel to the edge.

Stretched length. The bandage shall be laid on a horizontal surface and stretched to limit without application of mechanical force. While in the stretched position, the length of the bandage shall be measured parallel to the edges, to the nearest inch.

Elasticity test. A full width bandage (not previously stretched) shall be laid out on a flat horizontal surface without tension. A ten inch portion of the bandage shall be marked off and stretched to 22 inches. The bandage shall remain stretched to 22 inches for one hour, and then released. The marked off portion of the bandage shall be measured again (without tension). The percentage of stretch and return shall be computed as follows:

$$\text{Stretch, percent} = \frac{L_2 - L_1}{L_1} \times 100$$

$$\text{Return, percent} = \frac{L_2 - L_3}{L_1} \times 100$$

Where:

L_1 = Original length, prior to stretching.
L_2 = Stretched length.
L_3 = Relaxed length, after stretching.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:39Z
Check the source to verify that this is the current version before use.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

NOTES: The following National Stock Numbers are covered by this document:

NSN	Item Identification	Size
6510-00-935-5820	BANDAGE, ELASTIC, Rubber Warp Threads, 2 inches by 4-1/2 yards (5.08 cm by 41.1480 m), 12s.	1
6510-00-935-5821	BANDAGE, ELASTIC, Rubber Warp Threads, 3 inches by 4-1/2 yards (7.62 cm by 41.1480 m), 12s.	2
6510-00-935-5822	BANDAGE, ELASTIC, Rubber Warp Threads, 4 inches by 4-1/2 yards (10.16 cm by 41.1480 m), 12s.	3
6510-00-935-5823	BANDAGE, ELASTIC, Rubber Warp Threads, 6 inches by 4-1/2 yards (15.24 cm by 41.1480 m), 12s.	4

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6510-0811

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54874	21 May 1993

3. DOCUMENT TITLE
BANDAGE, ELEASTIC, RUBBER WARP THREADS

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON
c. ADDRESS (Include Zip Code)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198.290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.