INCH-POUND

A-A-54747
15 September 1992

COMMERCIAL ITEM DESCRIPTION

BANDAGE, FELT, ORTHOPEDIC, (RAYON)

The General Services Administration has authorized the use of this Commercial Item Description.

Salient Characteristics:

Shall be viscose rayon felt bandage suitable for padding under orthopedic casts. Bandages shall be of the following sizes:

SIZE	WIDTH (INCHES) +/- 1/4 INCH
1	3
2	4
3	6

Bandages shall have a minimum average length of 4 yards.

Bandages shall consist of the following materials and composition:

Shall be fully bleached, clean, dull, crimped viscose rayon or a mixture of 50 percent fully bleached, clean, dull, crimped, viscose rayon and 50 percent fully bleached, clean, bright, crimped, viscose rayon.

Not less than 70 percent, by weight, of fibers shall be 3/4 inch (19 mm) or greater in length; and not more than 5 percent, by weight, of the fibers shall be 1/4 inch (6 mm) or less in length. Fibers shall have a denier between 2.7 and 3.3. The rayon fibers shall be mechanically locked and entangled.

AMSC\NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Bandages shall meet the following requirements of Absorbent Gauze, USP:

Absorbency. The roll shall completely submerge in not more than 30 seconds.

Water extract (residue weight prior to ignition) shall not exceed 1.0 percent.

Residue on ignition shall not exceed 2.0 percent.

Acid or Alkali.

The material shall have a minimum weight of 60 grams per square yard when determined in accordance with method 5041 of Federal Test Method Standard No. 191. The average of two diameters of any roll shall be 2-1/8 to 2-3/4 inches. The average diameter of the roll shall be determined without compressing, by taking the visual minimum diameter, and then by taking the diameter at right angles to the minimum diameter. The finished surface of the finished material shall be either smooth or micro-pleated.

The bandages shall be in one continuous piece and neatly rolled. The finished bandages shall be capable of adhering to themselves and shall conform to the contour of the body or limb to which they are applied. Bandages shall not slip or roll up under plaster casts nor cause constriction nor have tourniquet effect on the body or limb to which they are applied.

Bandages shall be free of lumps, tears, holes and cuts. Bandages shall be substantially free of oil spots, streaks, dirt and foreign matter. Bandages shall not fray when applied nor during use.

Workmanship. The bandages shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG).

For Size 1, and Size 2, one package containing 72 bandages of one size only, as specified, constitutes one unit.

For Size 3, one package containing 36 bandages, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:38Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T09:38Z
Check the source to verify that this is the currentversion before use.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-00-817-2632	BANDAGE, FELT, ORTHOPEDIC, 6 inches by 4 yards, 36s	3
6510-00-817-2633	BANDAGE, FELT, ORTHOPEDIC, 3 inches by 4 yards, 72	1
6510-00-817-2634	BANDAGE, FELT, ORTHOPEDIC, 4 inches by 4 yards, 72s	2

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
USPHS
FDA-MPQAS

PGC 05206

Project No. 6510-0844

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54747	15 SEP 1992

3. DOCUMENT TITLE

BANDAGE, FELT, ORTHOPEDIC, (RAYON)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial (215)7372870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.