INCH-POUND

A-A-54756
30 September 1992

COMMERCIAL ITEM DESCRIPTION

BANDAGE, FELT, ORTHOPEDIC (COTTON)

The General Services Administration has authorized the use of this Commercial Item Description.

Salient Characteristics:

The finished bandage shall be suitable for padding under orthopedic casts, shall be capable of adhering to itself, shall conform to the contour of the body or limb to which applied and shall not slip or roll up under casts nor cause constriction of tourniquet effect. Bandages shall be of the following sizes:

SIZE	WIDTH (+/- 1/8 INCH)
1	2
2	3
3	4
4	6

The finished bandage shall have an unstretched length of four yards when made from the plain materiel and shall have a stretched length of four yards when made from the creped and compacted material. The finished bandage shall be one continuous piece.

Bandages shall be nonwoven cotton felt made from fully bleached purified new or reworked fine filament cotton. During fabricated cotton fibers shall be physically locked and entangled. Material shall meet the requirements for Purified Cotton, USP, for the following:

1. Residue on ignition, except that residue shall not exceed 0.50 percent

2. Water soluble substances, except that the residue shall not exceed 0.75 percent.

AMSC\NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

3. Absorbency.

4. The cotton from which the material is made shall meet the requirements of Purified Cotton, USP, for fatty matter.

The material shall be either plain or shall be creped and compacted. The plain material shall have a plain surface and shall have a minimum weight of 55 grams per square yard and a minimum average weight of 71 grams per square yard after fabrication of the bandage. The creped and compacted material shall have a rippled surface and shall have a minimum weight of 33 grams per square yard and a minimum average weight of 58 grams per square yard after fabrication of the bandage. The weights shall be measured by method 5041 of Federal Test Method Standard No. 191. Convert to grams per square yard. The material shall have a pH of 5.0 to 7.0 and shall be all cotton and shall not contain any binders or fillers.

Ply adhesion (shear strength). The plain material shall have a minimum average of 25 grams with no individual test result less than 18 gram and the creped and compacted material shall have a minimum average of 100 grams with no individual test result less than 90 grams when tested as follows. Condition samples for 24 hours at 70 $\pm$ 2° F and 50 $\pm$ 2 percent RH. Ten tests shall be performed comprising twenty samples cut 2 by 7 inches with the 7 inches cut in the machine direction. Superimpose one sample over the other so that there is a 5 inch overlap. Ply the samples together using a hand roller. Center the roller on the sample and slowly roll the overlapped area back and forth one time without exerting any added pressure when rolling. Only the weight of the 4.5 pound roller shall be on the sample. The hand operated roller shall meet the requirements of the Pressure Sensitive Tape Council Standards Appendage B. Immediately test samples after rolling by clamping the free end of each ply in the tensile machine jaws and pull (shear) the samples apart. The tensile machine shall be a constant-rate-of-extensiom machine using a load cell. The tensile machine shall be capable of measuring one gram and shall be accurate within $\pm$ 2%. The crosshead speed shall be 12 inches per minute.

The diameter of the finished roll shall be 2-1/8 $\pm$ 1/4 inch for the plain material and 2-3/4 $\pm$ 1/4 inch for the creped and compacted material. Rolls shall be measured at two places 90° from each other, then averaged.

Bandages shall be free of dirt, other foreign matter, lumps and cuts. Bandages shall be evenly rolled.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Workmanship. The bandages shall be free from their defects which detract from their appearance or impair their serviceability.

Unit. Package (PG).

For Sizes 1, 2 and 3, one package containing 72 bandages, as specified, constitutes one unit.

For Size 4, one package containing 36 bandages, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-00-133-4369	BANDAGE, FELT, Orthopedic, 4 inches by 4 yards, 72s	3
6510-00-133-4378	BANDAGE, FELT, Orthopedic, 2 inches by 4 yards, 72s	1
6510-00-200-3030	BANDAGE, FELT, Orthopedic, 3 inches by 4 yards, 72s	2
6510-00-200-3040	BANDAGE, FELT, Orthopedic, 6 inches by 4 yards, 36s	4

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
USPHS
FDA-MPQAS

PGC 05197

Project No. 6510-0840

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54756	30 SEP 1992

3. DOCUMENT TITLE

BANDAGE, FELT, ORTHOPEDIC (COTTON)

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial: (215) 7372870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 Previous editions are obsolete. 1992

Source: https://assist.dla.mil - Downloaded: 2026-02-26T09:38Z
Check the source to verify that this is the current version before use.