METRIC

A-A-54947
15 March 1994

COMMERCIAL ITEM DESCRIPTION

TUBE, TRACHEOSTOMY, Disposable, Jackson

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a disposable, Jackson, Size 6, tracheostomy tube used for long term ventilation of tracheostomy patients when dual cuff sites are required.

Salient characteristics:

General. The tracheostomy tube shall consist of the tube assembly with attached double cuff, neckpiece and dual inflation and relief valve assemblies. A double swivel ventilation attachment, neck tie cord, and disconnecting wedge shall also be provided.

Material. The tracheostomy tube and cuffs shall be fabricated of polyvinylchloride. Other components or parts except the tie cord shall be fabricated of suitable medical grade plastics. The tie cord shall be fabricated of cotton twill cord.

Style and design. The tracheal tube shall be in accordance with American National Standard Institute's (ANSI) standard for anesthesia equipment tracheal tubes, ANSI Z79.14. The outer cannula shall be a Jackson, Size 6 tracheostomy tube having and 8.0 millimeter (mm) inside diameter (ID), 11.0 mm outside diameter (OD), by 93 mm long. The cannula shall be radiopaque and shall have two thin walled, high volume low pressure cuffs. Each cuff shall be inflated by tubing

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

attached to a pilot balloon and relief valve. The inflation fitting shall be a luer fitting conforming to International Standards Organization (ISO)594. The relief valve shall limit cuff pressure to approximately 2.5 milligram (mg) of mercury (Hg). The cannula shall have an integral 15 mm connector and a winged neck flange that conforms to neck anatomy to minimize trauma due to patient neck motion. The size shall be clearly imprinted on the neck flange which shall also have slots for a tie cord. The 15 mm swivel connector shall be in accordance ANSI Z79.2 standard for tracheal tube connectors and adapters. The connector shall also be designed to allow suctioning without disconnecting the patients ventilator. A suitable tie cord 12.5 mm (1/2 inch) wide by 110 cm (48 inches) long shall be provided to secure the assembly to the patients neck. A plastic wedge shall be provided to allow easy disconnect of the swivel connector on ventilator.

Pressure test. The cuffs shall be inflated with air to a pressure of 25 mm Hg for a period of 30 minutes. There shall be no leakage or damage to the assembly.

Over pressure test. The pressure relief valves shall operate at a pressure of 28 ±2 mm of Hg.

Radiopacity. The insertion portion of the cannula shall be radiopaque when tested in accordance with American Society for Testing and Material (ASTM) Designation F640.

Sterility and pyrogenicity. The tracheostomy cannula assembly and all components shall be free from pyrogens when tested in accordance with the United States Pharmacopeia (USP). The item shall be sterile in accordance with the USP or American Association for Medical Instrumentation (AAMI) guidelines. Sterility assurance level shall be 0.000001.

Biocompatability. The item shall meet the requirements of the USP for Class IV.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Lead content. The tracheostomy tube parts in contact with the throat shall contain not more than 5 parts per million (ppm) of lead when tested in accordance with the USP.

Workmanship. The tracheostomy tube shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One tracheostomy tube, as specified, constitutes one unit. The tube shall be packaged in a suitable sealed container capable of maintaining sterility of content unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one tracheostomy tube. Sampling shall be inspection level S-2 with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one tracheostomy tube. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE 1: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-376-6558	TUBE, TRACHEOSTOMY, Disposable, Jackson, 8.0 mm ID by 11.0 mm OD by 93 mm long, Size 6, Sterile, Plastic

NOTE 2: ANSI and ISO standards are available from the American National Standards Institute Inc., 11 West 42nd Street, New York, NY 10036.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

A-A-54947

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5691
Location: tube/Intergraph/vab

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54947	2. DOCUMENT DATE (YYMMDD) 940315

3. DOCUMENT TITLE
TUBE, TRACHEOSTOMY, Disposable, Jackson

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial 215-737-2870 (2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.