INCH-POUND

A-A-54948
29 April 1994

COMMERCIAL ITEM DESCRIPTION

SUCTION APPARATUS, SURGICAL

(Gastrointestinal Abdominal Drainage)

(Programmable, Battery or AC Operated)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers three types of programmable, intermittent suction apparatuses for use during surgery.

Salient Characteristics:

The apparatus shall be in conformance with ASTM F 960 as applicable to surgical and ONT (Oral, Nasal, Tracheal) and gastrointestinal abdominal drainage suction apparatus.

Basic unit for Types I, II, and III. The unit shall be a programmable, intermittent suction apparatus for oral, nasal, and tracheal suctioning during surgery and post operative drainage. The apparatus shall be capable of operating from multiple power sources including a component battery.

The suction apparatus shall consist of at least a programmable, intermittent suction system including battery, charger circuit, controls, luminous indicators, a vacuum gauge, a suitable clear patient suction tube at least 5 feet long by 1/4 inch ID, two operator's manuals, two service manuals, two spare fuses for DC circuit (if fuse is used, otherwise a circuit breaker shall be incorporated into the DC circuit), one spare fuse for AC circuit (if fuse is used, otherwise a circuit breaker shall be incorporated into the AC circuit), and a disposable hydrophobic bacterial filter.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:40Z
Check the source to verify that this is the current version before use.

The suction apparatus shall be capable of operating using the following power sources:

- component battery
- external 12 VDC
- 115/230 volt, 50/60 Hz AC

The suction apparatus shall be capable of operating from any of the specified AC power sources while the internal battery is being charged via the component battery charging circuit.

The power cord shall be a minimum of 6 feet long.

The suction apparatus shall provide for a low and high range of vacuum.

In the low range of vacuum, the unit shall be capable of being user programmed to provide any of at least 144 intermittent timing cycles. The selection of ON time periods shall be from 5 seconds through 60 seconds in increments of 5 seconds. The OFF time period selection shall be the same as selection of ON time periods. The ON and OFF settings shall be independent of each other. When set to high range of vacuum, the unit shall provide the selected vacuum continuously. The ON/OFF timers shall be electronic.

The apparatus shall be so designed that it shall be capable of being operated a minimum of 300 hours over a time period of 6 months without requiring calibration.

The suction system shall provide for at least:

Controls:

- Master power switch
- Power mode switch(s) - external/internal/charge
- Vacuum mode switch - intermittent/constant
- Vacuum range switch - low/high
- Vacuum level switch - OFF through highest setting
- ON time selector - 5 through 60 sec. 5 sec. incr.
- OFF time selector - 5 through 60 sec. 5 sec. incr.
- DC voltage protector - fuse or circuit breaker
- AC voltage protector - fuse or circuit breaker
- Safety lockout - no intermittent in high range

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:40Z
Check the source to verify that this is the current version before use.

Switches may be electronic controls incorporating luminous indication of mode/position.

Indicators:

° Master power lamp *
° External electrical power and recharge lamp *
° Intermittent suction mode lamp *
° Low vacuum ON lamp *
° Vacuum gauge, dual scale, metric and English *

* A lamp may be an electronic read-out of condition or otherwise luminous indication of condition.

Minimum capabilities with patient tube and either the specified component collection system or the specified compatible canister system in place:

° Shall have a vacuum range of:

1. Operating on AC power, vacuum level range, selective from 0 to at least:

   a. High range, continuous vacuum, 550 mm Hg.
   b. Low range, continuous and intermittent, 200 mm Hg.

2. Operating on DC power, vacuum level range, selective from 0 to at least:

   a. High range, continuous vacuum, 550 mm Hg.
   b. Low range, continuous and intermittent, 120 mm Hg.

° High range free air flow, patient end of 5 ft. x 1/4 in. ID tube, selective range from 0 Lpm to at least:

1. With the hydrophobic bacterial filter in place:

   a. 18 Lpm when operated on AC power.
   b. 16 Lpm when operated on DC power.

2. Without the filter attached:

   a. 30 Lpm when operated on AC power.
   b. 25 Lpm when operated on DC power.

3. Low range, free air flow, patient end of 5 ft. x 1/4 in. ID tube, AC or DC power, with or without filter, selective within 10 percent, from 4 Lpm to 10 Lpm.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

° Battery operating time, 1 Hr., high range, 550 mm Hg
° Depleted battery charge time, max. 16 hr., 22 deg. C plus or minus 1 deg.
° Intermittent low range provide at least 144 settings
° Compatible with collection system defined in Type I
° Compatible with mobile cart defined in Type II

The basic unit shall be housed in a suitable case which shall not be larger then 0.85 cubic feet and the combined weight shall not exceed 28 pounds. The case shall be corrosion-resistant or treated for corrosion resistance. Any painted surfaces shall have at least one coat of commercial paint over a suitably prepared surface. The finish shall be free of gaps, scratches, chips, and other imperfections. The enclosure for electrical circuitry, including attachment of controls and indicators shall be drip proof in accordance with Underwriters Laboratories, Inc. (UL) 50, Type 5.

Type I. Type I shall include a collection system. The collection system shall include two collection canisters, a wall-mounting bracket system, and all required interconnecting tubing between canisters and the suction apparatus.

The design of the unit shall allow easy removal and installation of the internal battery without extensive dismantling of the unit. The time required for the procedure to remove or install the battery(ies) by a person experienced at this operation shall not exceed 5 minutes. The procedure shall start and end with a completely assembled unit (less battery(ies) as appropriate). The procedure shall not include disconnection/removal of mechanical/electromechanical components or disconnection of electrical circuits other than plug type connection for the battery(ies). The operating manual shall provide instructions for removing and installing the battery(ies).

The canisters shall have a volume of at least 1800 cc with a usable capacity of at least 1500 cc, capable of being autoclaved repeatedly at 250°F, and shall be clear glass or clear plastic. The canisters shall be graduated in increments of not more than 100 cc/ml from zero to, at least, 1500 cc/ml.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

The design of the mounting bracket and collection canisters shall allow easy removal of canisters for emptying. The bracket system shall be compatible with the disposable canister assembly described in Commercial Item Description (CID) A-A-51602 (latest issue as of date of this solicitation). In addition, the mounting bracket and collection system shall be compatible with the mobile cart described in Type II allowing attachment and detachment to/from the cart and alternately be attached to a wall.

Tubing shall include a 10 foot length by 3/8 inch ID tube between the suction pump and the filter, and interconnecting tubes of appropriate length by, at least 1/4 inch ID.

When the collection system is connected to the basic suction apparatus, the combined unit shall be a complete operating system for oral, nasal and tracheal suctioning.

The combined weight of the suction apparatus and collection system shall not exceed 50 pounds.

Type II. The Type II unit shall include a castered cart with a detachable mounting bracket system to secure the non-component collection system as described in Type I above.

The mounting bracket system shall be compatible with the disposable canister system described in Commercial Item Description (CID) A-A-51602 (latest issue as of date of this solicitation). The mounting bracket system shall be capable of being detached and wall mounted. The bracket shall be compatible with both the reusable collection system described in Type I and the disposable canister assembly described in CID A-A-51602.

The cart shall include, at least, 4 swivel casters, two of which shall be lockable. The casters shall be, at least, 4 inches OD.

The combined system of the basic suction apparatus and cart with mounting bracket shall be ready for mounting and connecting the collection system described in Type I or CID A-A-51602. The combined height shall be 40 inches (plus/minus 4 inches) and shall require a floor space of 18 inches (plus/minus 2 inches) by 18 inches (plus/minus 2 inches). The combined weight of the basic suction unit and cart with mounting bracket shall not exceed 67 pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Type III. The Type III unit shall include a collection system as defined in Type I and a cart as defined in Type II except only one mounting bracket system is required to secure either of the collection systems. The combined weight shall not exceed 77 pounds.

Workmanship. The suction apparatus shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One suction apparatus, as specified, constitutes one unit. The battery shall be packaged separately.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of I. The unit of product for examination purposes shall be one suction apparatus.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of I. The unit of product for test purposes shall be one suction apparatus.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:40Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasib specification requirem shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the applicable tests in ASTM F960.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Medical Procurement Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTES:

1. CID A-A-51602 covers NSN 6515-01-106-7648, Canister Assembly, Suction, Surgical. Commercial canisters that previously have met the salient characteristics of the CID are Baxter Healthcare (one outer canister 65651-115 and one Pkg of 100 liners 65652-551) and Abbott Laboratories (one outer canister 43423-04-01 and 2 Pkgs of 50 liners 43037-04-01).

2. The following National Stock Numbers are covered by this document.

6515-01-167-7287 - Type I
6515-01-180-6984 - Type II
6515-01-161-2938 - Type III

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5616
Location: SUCTN12C.DOC

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the currentversion before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. *Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.*

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54948	2. DOCUMENT DATE (YYMMDD) 940429

3. DOCUMENT TITLE SUCTION APPARATUS, SURGICAL (Gastrointestinal Abdominal Drainage) (Programmable, Battery or AC Operated)

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) 7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.