INCH-POUND

A-A-54953
13 May 1994

COMMERCIAL ITEM DESCRIPTION

# MANIKINS, CARDIOPULMONARY-RESUSCITATION, ADULT

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers two types of adult-size manikins.

Salient characteristics:

The manikin shall be an adult-size unit, suitable for teaching cardiopulmonary resuscitation (CPR) according to the requirement of the American Red Cross. Two types shall be offered:

Type I - full body
Type II - torso only

At a minimum, the manikin shall be a head section connected to a thorax section. The Xipoid and lower outline of the rib cage shall be palpable beneath the skin covering the torso section. A jersey or shirt shall be provided to cover the torso section.

Each manikin shall be a non-rebreathing manikin with a separate mouthpiece or facepiece for each student.

When the manikin is used in the artificial ventilation mode, it shall be necessary to tilt the head back to permit air to enter the lung. When artificial ventilation is performed properly, the chest shall visibly expand. The artificial lung and airway system shall be disposable or shall be capable of being sanitized, after each student's use. The entire airway system shall be easily disassembled.

AMSC N/A FSC 6910

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

When the manikin is used in the external cardiac compression mode, the sternum shall exhibit the same response as a human subject. Practicing of abdominal thrusts for the care of the unconscious choking victim shall be permitted by the device. The manikin shall also incorporate a carotid pulse simulator.

The manikin shall have a built-in real time signal device(s). The device shall show inflation volume, pulse beats, chest compression, and correct compression hand placement. The device shall be such that it can be viewed from all sides.

A recording device shall be provided which shall create a permanent record of the following: inflation volume, duration; and rate; compression depth and rate; correct hand placement; relationship between compressions and relaxation of hands; and carotid pulse beats.

Each manikin shall also include:

* A minimum of five disposable airway-with-lung combinations

* A minimum of five separate, disposable/reusable face-pieces (mouthpieces)

* A polyfoam kneeling pad

* A sanitizing kit, including a measuring cup and pan with adequate instructions for proper sanitizing of manikins using bleach and water solution (or the latest government sanitizing standards)

* A carrying case large enough to contain the manikin and all other items included with the manikin

The following supplies shall be made available for purchase as required in the quantities specified:

* At least ten disposable/reusable facepieces

* One hundred (100) disposable airway/lung bags

* Recorder paper and other supplies

All manufacturer's user's literature shall include a complete parts breakdown with part numbers.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

The adult manikin shall be capable of operating over the full tolerance range of one or more of the following combinations of line voltages and frequency combinations (unless battery operated) as specified by the customer requisition.

* 115 VAC plus or minus 10 percent, 50 plus or minus 2 Hz.
* 115 VAC plus or minus 10 percent, 60 plus or minus 2 Hz.
* 230 VAC plus or minus 10 percent, 50 plus or minus 2 Hz.
* 230 VAC plus or minus 10 percent, 60 plus or minus 2 Hz.

A dual voltage and hertz feature on all units is desirable, but not mandatory.

Change-over devices, switches, circuits, or transformers, if used for the purpose of voltage or frequency conversion, shall be clearly marked and identified in the appropriate documentation which accompanies the unit. Mechanical or electrical adjustments to effect the conversion shall not require the use of specialized tools or skills.

All elements shall be certified to consist of nontoxic materials which will not degrade to a toxic state.

Workmanship. The adult cardiopulmonary resuscitation manikin shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One manikin, together with manufacturer's user's literature, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one manikin.

Sampling for tests. Sampling for tests, including dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one manikin.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

1. Operate signal device while using the manikin to confirm that proper ventilation and compression are indicted as specified.

2. Operate recorder while using the manikin to confirm that all parameters specified are correctly documented on the chart.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

A-A-54953

NOTE: The following National Stock Numbers are covered by this document.

6910-01-377-1430 - Type I
6910-01-377-1435 - Type II

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6910-0124
Location: MANK2C1D.DOC



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

| I RECOMMEND A CHANGE: | 1. DOCUMENT NUMBER | 2. DOCUMENT DATE (YYMMDD) |
|---|---|---|
| | A-A-54953 | 940513 |

3. DOCUMENT TITLE

MANIKINS, CARDIOPULMONARY-RESUSCITATION, ADULT

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

| a. NAME (Last, First, Middle Initial) | b. ORGANIZATION | |
|---|---|---|
| c. ADDRESS (Include Zip Code) | d. TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON (If applicable) | 7. DATE SUBMITTED (YYMMDD) |

8. PREPARING ACTIVITY

| a. NAME | b. TELEPHONE (Include Area Code) | |
|---|---|---|
| Defense Personnel Support Center<br>ATTN: DPSC-MSE | (1) Commercial<br>215-737-2870 | (2) AUTOVON<br>444-2870 |
| c. ADDRESS (Include Zip Code)<br>2800 South 20th Street<br>Philadelphia, PA 19145 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466<br>Telephone (703) 756-2340 AUTOVON 289-2340 | |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26109:40Z
Check the source to verify that this is the current version before use.