INCH-POUND
A-A-54957
10 June 1994

COMMERCIAL ITEM DESCRIPTION

BAG, OSTOMY

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers stoma bags suitable for post-operative wear over an artificial opening related to the colon or ileum for containing and disposing of ostomy discharge.

Salient characteristics:

Classification. The ostomy bags shall be furnished in the following sizes:

Size	Inside Diameter (I.D.)
1	2 inches
2	2-1/2 inches

Material. The bag shall be fabricated of polyethylene. The Karaya seal shall consist of karaya powder, glycerine, and mold and bacteria inhibitors. The clamp bottom closure shall be fabricated of polypropylene. Bag material shall be non-toxic and non-irritating to body skin surfaces.

Style and design. The bag shall be a drainable disposable stoma bag of medium length with the body fabricated of co-extruded polyethylene. The polyethylene shall have a minimum thickness of 2 mils. The bag shall have approximate dimensions of 15-1/2 inches long by 6-1/2 inches width at the top and 2-1/2 inches at the bottom. The bottom edge of the bag shall be open to permit drainage of contents. The bag shall be provided with a polyethylene gasket with a Karaya seal attached on the top center back surface for

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

sealing to the body. The gasket shall have a tab, with a hole, on each side for attachment of a ostomy bag belt. A clamp fabricated of polypropylene shall be provided to seal the bottom opening. The clamp shall be one piece, having a smooth surface with an integral hinge and easy-open latch. The clamp length shall be approximately 3-1/4 inches.

The bag component dimensions shall be as follows:

	Gasket (Inches)			Seal (Inches)		
Size	I.D.	O.D.	Length	I.D.	O.D.	Thickness
1	2	2-3/4	4-1/2	1-5/8	2-3/4	13/64
2	2-1/2	3-1/4	5-1/4	2-1/8	3-1/4	13/64

Leakage. Bags shall be completely filled with cold tap water. The bags shall not leak when suspended from a suitable support for eight (8)hours.

Instructions. Each ostomy package shall contain instructions for use.

Workmanship. The bags shall be free from defects which detract from their appearance or may impair their serviceability. Bags shall be free from tears, cuts, weak areas, foreign material or other defects.

Identification marking. Each bag shall be legibly and permanently marked with the manufacture's name or registered trademark.

Unit. Package (PG). One package containing 30 bags of one size only, together with instruction for use, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one package of ostomy bags. Sampling shall be inspection level S-2 with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one package of ostomy bags. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications and Sizes are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6515-00-915-4143	BAG, OSTOMY, Karaya Seal, Drainable, Disposable, 2 inch size, 30s	1
6515-00-828-2469	BAG, OSTOMY, Karaya Seal, Drainable, Disposable, 2-1/2 inch size, 30s	2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

A-A-54957

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5716
Location: bag/Intergraph/vb

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54957	10 June 1994

3. DOCUMENT TITLE
BAG, OSTOMY

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial (215)7372870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the current version before use.