METRIC

A-A-54961
30 June 1994

COMMERCIAL ITEM DESCRIPTION

INTRAVENOUS INJECTION SET

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an intravenous injection set suitable for accurate administration of parental solutions at 10 drops per milliliter (ml).

Salient characteristics:

Material. The tubing shall be fabricated from PVC and the adapters, cap, and clamp shall be fabricated from rigid, dimensionally stable plastic. The tubing shall be transparent, flexible and pliable, but sufficiently rigid to prevent collapse during normal use. The tubing shall be free from all liquid additive substances which oxidize, exude or sweat-out onto the surfaces of the tubing.

Style, design and dimension. The intravenous injection set shall be a non-vented, sterile, non-toxic, disposable, closed system intravenous infusion device having a cap with piercing device, a flexible plastic calibrated drip chamber, plastic tubing at least 250 centimeter (cm) (100 inches) long, two "Y" injection sites, a male needle adapter, a roller flow control clamp, a .22 micron air eliminating filter, check valve and end protectors.

One "Y" injection site shall be approximately 15.24 cm (6 inches) from the male needle adapter and the other "Y" injection site shall be 202.50 cm (81 inches) from the male needle adapter. The set shall be shipped with the clamp in the open position.

The injection set shall be compatible with all flexible and semi-rigid parental solution containers. Luer fittings shall be in accordance with International Standards Organization (ISO) 594/1.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Safety (non-toxicity) and pyrogens. The injection set shall be nontoxic (safe) and pyrogen-free when tested in accordance with the United States Pharmacopeia (USP) for transfusion and infusion assemblies.

Sterility. The injection set shall be sterile in accordance with the USP or the Association for the Advancement of Medical Instrumentation (AAMI) requirements. Sterility assurance level shall be 0.000001.

Biologics. The injection set shall meet the requirements for Class II plastics stated in the USP.

Deformation of tubing. While held in the same position on the tubing, the flow regulator shall be operated from its fully open position to complete shut-off ten times. The tubing shall not be cut, pierced, or damaged in any manner which would cause leakage or restriction of the flow of solution.

Diaphragm reseal test. The rubber diaphragm shall reseal without a falling drop within 10 seconds after puncture with a dry nonlubricated 18 gauge needle when subjected to an internal (hydrostatic) pressure of 4 psig for 60 seconds.

Instructions. Instructions for proper usage of the sets shall be provided with each unit package.

Workmanship. The intravenous injection sets shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 48 individually packaged injection sets, as specified, constitutes one unit. Each set shall be individually packaged in a suitable sealed container capable of maintaining sterility of contents unless package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the current version before use.

set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one intravenous injection set as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one intravenous injection set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:40Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-082-1021	INTRAVENOUS INJECTION SET, Non-Vented, With Flexible Drip Chamber, Without Needle, Disposable, 48s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5723
Location: c: nish\ivinjset

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54961	30 June 1994

3. DOCUMENT TITLE
INTRAVENOUS INJECTION SET

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial (215)7372870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.