METRIC

A-A-54962
30 June 1994

COMMERCIAL ITEM DESCRIPTION

INTRAVENOUS INJECTION SET, Secondary

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a secondary intravenous injection set for use with primary piggyback sets.

Salient characteristics:

The intravenous injection set shall be a vented, sterile disposable set measuring approximately 81 centimeter (cm) (32 inches) from the base of the spike to the male Luer fitting. The set shall have a spike with a bacterial fitter air vent connected to a flexible drip chamber, a slide clamp to regulate fluid rate, an 18 or 19 gauge needle suitable for intravenous therapy, a male Luer fitting and interconnecting I.V. plastic tubing.

The drip chamber shall be calibrated to a nominal 15 drops/ milliliter (ml). The tubing inside diameter shall be 0.250 cm and have a priming capacity of 6 ml. The tubing and drip chamber shall be clear and transparent. The set shall include an extension hook. The set shall be compatible with flexible plastic and glass containers with solid rubber stoppers. Protective covers shall seal the needle and fluid path. Tubing shall be free from all liquid additive substances which oxidize, exude or sweat-out onto the surfaces of the tubing. The Luer fitting taper shall be in accordance with International Standards Organization (ISO) 594/1.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Safety (non-toxicity) and freedom from pyrogens. The fluid path shall be nontoxic (safe) and pyrogen-free when tested in accordance with the United States Pharmacopeia (USP) for Transfusion and Infusion assemblies.

Biologics. The plastics shall conform to the requirements of USP Class II plastics.

Sterility. The set shall be sterile in accordance with the USP or AAMI guidelines for Transfusion and Infusion assemblies. Sterility assurance level shall be 0.000001.

Workmanship. The intravenous injection set shall be free from defects which detract from its appearance or impair its serviceability. Components shall be free from cuts, abrasion, discolored areas, blisters and contamination.

Unit. Package (PG). One package containing 48 individually packaged intravenous injection sets, as specified, constitutes one unit. Each set shall be supplied in a suitable sealed container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one intravenous injection set. Sampling shall be inspection level S-2 with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one intravenous injection set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-107-1201	INTRAVENOUS INJECTION SET, Secondary, with Flexible Drip Chamber, Vented, 18 or 19 gauge needle, Disposable, 48s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5721
Location: Intravenous/Intergraph/vb

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. *The preparing activity must complete blocks 1, 2, 3, and 8* In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. *The preparing activity must provide a reply within 30 days from receipt of the form.*

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54962	2. DOCUMENT DATE *(YYMMDD)* 30 June 1994

3. DOCUMENT TITLE
INTRAVENOUS INJECTION SET, Secondary

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME *(Last, First, Middle Initial)*	b. ORGANIZATION	
c. ADDRESS *(Include Zip Code)*	d. TELEPHONE *(Include Area Code)* (1) Commercial (2) AUTOVON *(If applicable)*	7. DATE SUBMITTED *(YYMMDD)*

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center ATTN: DPSC-MSE	b. TELEPHONE *(Include Area Code)* (1) Commercial: (215)7372870 (2) AUTOVON: 444-2870
c. ADDRESS *(Include Zip Code)* 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.