METRIC

A-A-54963

30 June 1994

COMMERCIAL ITEM DESCRIPTION

IMPLANT, ORBITAL, SPHERE (MULES)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers ophthalmic eye-sphere implants for use in the eye socket.

Salient characteristics:

Style and design. Shall be spherical ophthalmic eye implants. The diameter of each implant covered by this document shall be as follows:

Type	Diameter (millimeters)
1	16 ± 0.5
2	18 ± 0.5
3	20 ± 0.5

Each implant listed above shall be supplied with a universal conformer. The conformer shall be designed for suture fixation of the eye sphere.

Material. The implants and conformer shall be fabricated from non-rigid, surgical-grade silicone rubber. This material shall comply with all requirements as specified below. Samples of this material shall be subjected to test specified for each requirement.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Requirements for "Surgical - Grade Silicone Rubber"

Requirement		Test Method
Specific Gravity:	1.12 - 1.25	D792
Shore Hardness:	A45 - A55	D22240
Water Absorption:	0.05 w/o, maximum	D570(1)
Elastic Modulus:	10,000 psi, maximum	D638(2)
Toxicity:	Non-Toxic	F895(3)

(1) Perform the 24 hour water absorption test.
(2) Perform this test at 23°C (73°F) and 50 percent relative humidity.
(3) As an alternate, perform the Class IV United States Pharmacopeia (USP) systemic injection test specified in the USP under "Biological Tests - Plastics".

Sterility. Sample implants and conformers shall be tested for sterility in accordance with the latest edition of the USP or the American Association for Medical Instrument-ation (AAMI) guidelines. Sterility assurance level shall be 0.000001.

Finish. The implants and conformer shall have a smooth, uniform finish.

Workmanship. The implants and conformer shall be free from blooming, blisters fish eyes, frosting, shorts, pits, pimples, shrink-marks, and any other defects that detract from their appearance or impair their serviceability. Definitions of defects listed above shall be in accordance with American Society for Testing and Material (ASTM) Designation D883, "Standard Definitions of Terms Relating to Plastics".

Unit. Each (EA). One eye implant and one conformer, as specified, constitutes one unit. Each unit shall be individually packaged in a sealed peel-open type container capable of maintaining sterility of contents unless package is damaged or opened.

Autoclaving. The implants and conformer shall be capable of steam sterilization at 250°F for a period of 30 minutes with no damage to these components.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one implant and conformer. Sampling shall be inspection level S-2 with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one implant and one conformer. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers, Types, and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	TYPE	ITEM IDENTIFICATION
6515-01-139-7606	1	IMPLANT, ORBITAL, SPHERE, Soft, Autoclavable, 16 mm
6515-01-139-7607	2	IMPLANT, ORBITAL, SPHERE, Soft, Autoclavable, 18 mm
6515-01-139-7608	3	IMPLANT, ORBITAL, SPHERE, Soft, Autoclavable, 20 mm

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5710
Location: implantorb/Intergraph/vb

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54963	30 June 1994

3. DOCUMENT TITLE

IMPLANT, ORBITAL, SPHERE (MULES)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial (215)7372870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.