INCH-POUND

A-A-54964
15 July 1994

COMMERCIAL ITEM DESCRIPTION

MONITORING SET, ARTERIAL BLOOD GAS SAMPLING AND PRESSURE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers the requirements for a sterile, pyrogen free, disposable arterial blood gas and pressure monitoring set suitable for use in operating rooms and intensive care units.

Salient characteristics:

Materials. All tubings shall be fabricated from clear, pliant polyvinychloride, suitable for pressure transmission from the patient to the transducer. The stopcocks and the continuous flush device shall be made of clear polycarbonate. The tubing, the stopcocks and the continuous flush device shall meet the requirements of the United States Pharmacopeia (USP) for Class IV plastics. The nipples shall be fabricated from suitable plastic. The monitoring set shall be sterile and pyrogen free and shall meet the sterility requirements when tested in accordance with the USP

Construction. The monitoring set shall consist of:

( ) A 12 inch (30.5 cm) pressure tubing with male/female Luer lock fittings on each end.

( ) A 48 inch (121.9 cm) pressure tubing with male/female Luer lock fittings on each end

( ) A 6 inch (15.25 cm) pressure tubing incorporating a continuous flush attachment system with male/female Luer lock fittings on each end

(4) Three (3), three way stopcocks

(5) Three (3) extra nipples.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

All tubings shall be kink resistant and have a nominal internal diameter of 0.065 inch (1.6 mm).

The stopcocks handle shall be marked for easy identification of the direction of flow. The handle action shall be smooth for ease of switching the flow direction

The continuous flush device shall provide continuous flow rate from 2 to 4 milliliters/hour

All tubings shall have nipples on both ends to prevent contamination during handling

All male/female Luer lock connections shall be in accordance with the International Standards Organization (ISO) 594

Workmanship The monitoring set shall be free from defects which detract from its appearance or impair its serviceability.

Unit Package (PG). Each monitoring set shall be individually packaged in a sealed peel-open container capable of maintaining sterility of contents unless package is damaged or opened One package containing 25 individually packaged monitoring sets, as specified, constitutes one unit

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one package of monitoring sets. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified Unit of product for test purposes shall be one package of monitoring sets. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product

Regulatory requirements.

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-125-3240	MONITORING SET, ARTERIAL BLOOD GAS SAMPLING AND PRESSURE, Plastic, Disposable, 25s

NOTE 2: ANSI and ISO Standards are available from the American National Standards Institute Inc.; 11 West 42nd Street; New York, NY; 10036.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5713
Location: tony/gas/dawne

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54964	2 DOCUMENT DATE (YYMMDD) 940715

3 DOCUMENT TITLE
MONITORING SET, ARTERIAL BLOOD GAS SAMPLING AND PRESSURE

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN DPSC-MSH	b TELEPHONE (Include Area Code) (1) Commercial (215)7372870	(2) AUTOVON 444-2870
ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041 3466 Telephone (703) 756 2340 AUTOVON 289 2340	

DD Form 1426, OCT 89 Previous editions are obsolete 198/290

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:40Z
Check the source to verify that this is the current version before use.