INCH-POUND

A-A-54968
31 July 1994

COMMERCIAL ITEM DESCRIPTION

IMPRESSION MATERIAL, DENTAL, (Tissue Conditioner Type)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a non-toxic impression and tissue treatment material suitable for allowing abused and distorted oral tissue to regain its normal anatomical configuration and physiological tone.

Salient characteristics:

Material. Materials shall be of the highest quality throughout and suitable in all respects for the purpose intended. Any ingredients entering into the impression material which are not included in "The United States Pharmacopeia" (USP) or "National Formulary" (NF) shall be of a quality suitable for pharmaceutical use.

The impression material shall conform to either one of the following two compositions:

Composition A. Powder shall be ethyl methacrylate resin. Liquid shall be denatured ethyl alcohol, and methyl salicylate and glycolate (plasticizers). Separating medium (lubricant) shall be essentially silicone, with the main ingredient being dimethylpolysiloxane.

Composition B. Powder shall be ethyl methacrylate resin and zinc undecylenate. Liquid shall be formulated of an aromatic ester, a fixed oil, and an ester plasticizer, and a small amount of ethyl alcohol and deodorant. A suitable lubricant for separating the impression material from the gypsum casts shall be furnished.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Description. The impression material shall be a non-toxic impression and tissue treatment material. The impression material shall consist of a bottle of powder, a bottle of liquid, a bottle of separating medium (lubricant), a powder measure, a liquid measure, a minimum of 12 disposable mixing cups (or one reusable mixing vessel), and instructions for use. A spatula and patient education sheet may also be furnished at the manufacturer's option.

The bottle of powder shall contain a minimum of 6 ounces. The bottle of liquid shall contain a minimum of 6 fluid ounces The volume of liquid furnished shall be 5 percent in excess of the total powder-liquid ratio required by the manufacturer's instructions. The bottle of separating medium shall contain a minimum of 1/2 fluid ounce. The powder and liquid measures and mixing cups shall be of a capacity sufficient for measuring and mixing individual portions. If a reusable mixing vessel is supplied, it shall have a capacity of 2-1/2 fluid ounces. The immediate containers and all components shall be compatible, and shall not interact physically or chemically with the powder, liquid, or mixed impression material.

Physical properties.

Odor and flavor. The material shall not have an unpleasant odor or flavor.

Fineness. Ninety-five to ninety-nine percent of the powder shall pass through an 80 mesh sieve.

Mixing time and uniformity. The time of mixing the impression material in accordance with the manufacturer's directions to obtain the required consistency shall be no more than one minute. The mix obtained shall be smooth, homogeneous, workable, and free from lumps and granules.

Pliability. When mixed according to manufacturer's instructions, impression material shall remain pliable, that is, shall be capable of deformation under an approximate pressure of 85 millimeters (mm) of mercury (the approximate equivalent of diastolic blood pressure), for a minimum of 2 days when applied inside provisional dentures.

Working time. The working time (time during which material is plastic and moldable) shall be no less than 3 minutes.

Hardness. Impression material shall have a Shore durometer hardness of between 5 and 12.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Amount of material in unit container. A single mix of the impression material, when prepared using the powder and liquid measures provided by the manufacturer, shall have a bulk displacement of between 10.8 and 15.2 milliliters (ml). The total amount of material in the unit of issue shall be sufficient to prepare a minimum of 25 impressions.

Instructions for use. Adequate and accurate instructions for manipulating the material shall accompany each unit container. These instructions must include the following as a minimum:

1. Preparation of the denture before treatment.
2. Preparation of the patient's mouth before treatment.
3. Proportions of powder and liquid to be used to give proper mix in terms of scoops of powder to scoops of liquid.
4. Method of mixing.
5. Mixing time.
6. Proper use and application of the lubricant (separating compound).

Instructions for use shall be furnished as a separate sheet in the unit container.

Workmanship. The impression material shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one bottle of powder, one bottle of liquid, one bottle of separating medium (lubricant), one powder measure, one liquid measure and either 12 disposable mixing cups or one reusable mixing vessel, together with instructions for use, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one package of impression material, as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one package of impression material. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-00-104-7617	IMPRESSION MATERIAL, DENTAL, Tissue Conditioner Type

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6520-2559
Location: paul/impress/dawne

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

   The submitter of this form must complete blocks 4, 5, 6, and 7.

2. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54968	2. DOCUMENT DATE (YYMMDD): 940731

3. DOCUMENT TITLE: IMPRESSION MATERIAL, DENTAL (Tissue Conditioner Type)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MAP	(1) Commercial (215) 737-8431	(2) AUTOVON 444-8431
ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.