INCH-POUND

A-A-54970
29 July 1994

COMMERCIAL ITEM DESCRIPTION

MICROSCOPE, OPTICAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a lightweight clinical microscope with external power supply, suitable for field use.

Salient characteristics:

The lightweight field microscope shall be usable, transportable and storable in hot, basic and cold climate with a temperature range of 90°F to -50°F.

The optical microscope shall weigh no more than 8 pounds with the mirror or lamp assembly and shall not exceed 10 inches in height without the binocular head. The overall dimensions for the microscope and separate power supply shall not exceed 13-1/2 x 11-1/4 x 6 inches.

The microscope shall be supplied with a suitably sized carrying case constructed of a high density polyethylene plastic. The carrying case shall include a handle, full length piano hinge and at least two twist-type latches. The carrying case shall include an air release valve, and shall be designed and fitted to provide an air and water-tight closure. A styrofoam insert which shall securely house the microscope assembly and power supply shall be provided with the case. Shall be olive drab in color.

AMSC N/A FSC 6650

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

The microscope shall be constructed of corrosion resistant material and shall consist of the following components:

* Microscope stand.

* Binocular head with dual diopter tubes and interpupillary adjustments. Head shall have inspection plate for easy maintenance. Removal of head shall be by knurled knob.

* Dual coarse and fine adjustment knobs. Coarse adjustment shall have two internal tension adjustment points.

* Attachable mechanical stage with coaxial controls, low positioned on ball bearing slideways.

* Mirror and fork assembly.

* Paired 10X DIN widefield eyepiece holding:

  4X DIN N. A. 0.10 objective
  10X DIN N. A. 0.25 objective
  40X DIN N. A. 0.65 retractable objective, capable of accepting 30 degree contrast plate.

* 100X DIN research grade oil immersion modified to accept darkfield funnel that matches darkfield stop

* Combination Abbe and darkfield condenser, N.A. 1.25 with iris diaphragm in spiral mount

* Differential illumination system; attachable, 6 volt, 10 watt krypton lamp, condensing lens and blue filter

* Knurled knob for replacement of bulb

* Instruction manual

The illuminator power supply shall have nominal dimensions of 4-3/4 x 4-3/4 x 2-1/3 inches with a single chip circuit board, and shall work on 6, 12, 24 volt, 110/220 VAC. Cables supplied with the illuminator shall allow the microscope to work off 24 volt tactical vehicle battery.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

The illuminator socket shall accept a double contact bayonet tungsten bulb. The bulb shall operate below maximum voltage, giving nominal bulb life in excess of 10,000 hours. Removal of the socket shall be by knurled knob.

A lockout ground plug or 3-to-2 plug (cheater), to be used where three-hole receptacles are not available, shall be included.

The power supply shall have a rechargeable 6 volt gel cell, with an automatic internal reset mechanism.

The entire unit shall require no tools for assembly, tear down, replacement of bulbs, etc. If required, special tools shall be supplied with the microscope.

All screws/knurled knobs shall be removable by gloved hands.

The microscope shall be nuclear, biological and chemical decontaminable using standard decontaminates.

Workmanship. The optical microscope shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One optical microscope, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an Acceptable Quality Level (AQL) of 2.5 (percent defective) and an inspection level of S-2. The unit of product for examination purposes shall be one microscope.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one microscope.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTES: 1. NSN: 6545-00-926-9037, Case, Centrifuge, Microscope, Reagents, is a suitable carrying case for the microscope in this specification.

2. The following National Stock Number is covered by this document.

6650-01-259-3008

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6650-0188

Location: MICOP14C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54970	940729

3 DOCUMENT TITLE
MICROSCOPE, OPTICAL

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a. NAME (Last, First, [illegible])	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) 7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code)
DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MAP	(1) Commercial (215)737-8431 (2) AUTOVON 444-8431
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041 3466 Telephone (703) 756 2340 AUTOVON 289-2340

DD Form 1426, OCT 89 Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.