METRIC

A-A-54978
31 August 1994

COMMERCIAL ITEM DESCRIPTION

CANNULA, TRACHEOSTOMY, Nylon

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers molded nylon tracheostomy cannulae used to provide a continuous airway when inserted in a surgical tracheostomy opening, and to maintain the opening.

Salient characteristics:

Material. Each component of the cannula shall be molded of high quality, natural colored, hydrolysis resistant, translucent nylon.

Style, design and construction. Shall be a lightweight, natural colored, translucent nylon, tracheostomy cannula. Shall consist of two concentric tubes, one fitting inside the other with adapter and an obturator for insertion of the outer tube. The curved outer tube shall be equipped with an integrally molded neck plate. The arc between the outside of the neck plate and the distal end of the tube shall be 90°. The inside diameter shall be uniform from the neck plate to a point 4.7 mm (3/16 inch) short of the distal end of the tube where it shall decrease sufficiently to form a contact fit with the outside diameter of the inner tube. The neck plate shall be elliptical and shall be shaped to conform to the contour of the average neck. It shall be provided with two curved transverse slots to accept tie tapes for securing cannula to the patient. The curved inner tube shall be equipped with an integrally molded 15 mm adapter conforming to ANSI Z79.2. The outside diameter of the inner tube shall form a contact fit at the distal end of the outer tube and when so fitted shall insert without drag. When fully inserted it shall extend slightly beyond the distal end of the outer tube. The obturator shall be molded integrally.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:41Z
Check the source to verify that this is the current version before use.

The proximal end shall be circular and cone-shaped to seat in the neck plate of the outer tube and shall provide a finger grip for removal. The distal end shall be bullet shaped and shall extend beyond the distal end of the outer tube to form a pilot for the introduction of the outer tube. It shall be so fitted to insert into and remove from the outer tube without drag. Wherever components are required to be integrally molded, they may be attached by means of ultrasonic welding.

The outer cannula dimensions shall be as follows:

SIZE	O. D. (mm)	Length (mm)
3	7	70
4	8	76
6	10	89
7	11	95
8	12	102

Tube lengths shall be measured on the outer curve from the distal end to the neck plate.

Sterilization. Cannulae and obturators shall not become distorted, cracked, or be adversely affected as to impair their use, when subjected to autoclave sterilization for 20 minutes at 137.80 kPa (20 psi) steam pressure.

Interchangeability test. Samples (inner cannula, outer cannula and obturator) shall be randomly selected, and shall be tested as follows: Inner and outer cannulae and obturators shall be completely separated. The inner and outer cannulae shall be reassembled, making certain that no inner cannula is mated with the same outer cannula from which it had been removed. The inner cannula shall be capable of moving in and out of the outer cannula without evidence of objectionable sticking and shall seat completely in the outer cannula. Obturators shall be inserted in outer cannulae, making certain that no obturator is mated from the same set from which it had been removed. Obturators shall fit outer cannulae.

Biocompatability. The cannula shall meet the requirements of Type II plastics when tested in accordance with the United States Pharmacopeia (USP).

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:41Z
Check the source to verify that this is the current version before use.

Identification marking. The tracheostomy cannula shall be permanently and legibly marked with the contractor's name or registered trademark. Components (outer tube and obturator) shall have a 3 mm raised molded number, designating size, at proximal end of each component.

Workmanship. Components shall be smooth finished, and shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One cannula, as specified, constitutes one unit. Cannula shall be supplied in a suitable sealed container.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one cannula as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one cannula. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:41Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers, their respective Item Identifications and their respective Sizes are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6515-00-914-0244	CANNULA, TRACHEOSTOMY, Nylon	3
6515-00-914-0245	CANNULA, TRACHEOSTOMY, Nylon	4
6515-00-914-0248	CANNULA, TRACHEOSTOMY, Nylon	6
6515-00-914-0249	CANNULA, TRACHEOSTOMY, Nylon	7
6515-00-914-0250	CANNULA, TRACHEOSTOMY, Nylon	8

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:41Z
Check the source to verify that this is the current version before use.

A-A-54978

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5725
Location: c: nish\cannula2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE: AA	1. DOCUMENT NUMBER A-A-54978	2. DOCUMENT DATE (YYMMDD) 940831

3. DOCUMENT TITLE
CANNULA, TRACHEOSTOMY, Nylon

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center ATTN: DPSC-MAP	b. TELEPHONE (Include Area Code) (1) Commercial: (215) 737-8431 (2) AUTOVON: 444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete. — 198-290

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:41Z
Check the source to verify that this is the current version before use.