INCH-POUND

A-A-54979
31 August 1994

COMMERCIAL ITEM DESCRIPTION

INCUBATOR, BACTERIOLOGICAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a dual chambered bacteriological incubator.

Salient Characteristics:

Shall be a dual vertical carbon dioxide incubator. Shall be suitable for wet or dry, and aerobic or CO_2 incubations of microbiological cultures.

Shall have each chamber independently regulated to accommodate microbiological cultures which require different and varying environments. Each chamber shall be provided with facilities for heating, humidity and CO_2 tension.

Shall have corrosion resistant steel chamber interiors with an exterior steel housing that is suitably protected from corrosion.

Shall have nominal dimension (inches) for each chamber of 17 width x 17 depth x 22 height.

The nominal outside dimensions (inches) of the overall housing shall be 24 width x 24 depth x 65 height.

Shall have the following operation capabilities for each chamber:

a. An adjustable temperature of 5° Centigrade (C) above ambient to 50°C.

b. A humidity range of ambient to 97 percent R.H.

AMSC N/A FSC 6640
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

c. Shall have adjustable CO_2 tension (automatic or manual) from 0 to 20 percent.

Shall have a thermostatic control for each chamber that maintains a temperature stability at any temperature setting of $\pm$ 0.5° C. A safety thermostat shall also be provided with each thermostat control to prevent excessive temperature build-up in the event that the primary thermostatic control should fail.

Each chamber shall be supplied with the following:

a. A minimum of five removable corrosion resisting shelves for each chamber.

b. A hinged inner door with a latch, that in closure makes a positive seal and allows for visibility of the chamber contents without disturbing the temperature equilibrium.

Shall have a two-stage regulation with 320 CGA fitting.

Shall have a suitable means (e.g. a dial thermometer) for visually checking the chamber temperature without having to open the outer door.

Shall have probe access ports which permit checking the CO_2 concentration in each chamber without the loss of tension.

Workmanship. The bacteriological incubator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One incubator, as specified, constitutes one unit. The incubator shall be supplied in a suitable unit carton.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one incubator.

Inspections and testing. Inspections and testing shall be conducted on the incubator in accordance with commercial practice. Sampling for the inspections and testing shall be in accordance with commercial practice. In addition, each incubator shall be tested to demonstrate its capability for proper operation.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document.

6640-01-187-6621

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6640-1861

Location: INCUB14C.DOC

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:40Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:

1 DOCUMENT NUMBER: A-A-54979

2 DOCUMENT DATE (YYMMDD): 940831

3 DOCUMENT TITLE: INCUBATOR, BACTERIOLOGICAL

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)

b. ORGANIZATION

c. ADDRESS (Include Zip Code)

d. TELEPHONE (Include Area Code)
(1) Commercial
(2) AUTOVON (If applicable)

7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME: DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DPSC-MAP

b TELEPHONE (Include Area Code)
(1) Commercial: (215)737-8431
(2) AUTOVON: 444-8431

c ADDRESS (Include Zip Code)
2800 South 20th Street
Philadelphia, PA 19145

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT
Defense Quality and Standardization Office
5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466
Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:40Z
Check the source to verify that this is the current version before use.