INCH-POUND

A-A-54981
20 January 1995

COMMERCIAL ITEM DESCRIPTION

HUMIDIFIER, PORTABLE

(Therapeutic)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an electrically operated portable humidifier.

Salient characteristics:

The humidifier shall consist of a plastic housing (medium or high impact molded polystyrene, polypropylene, or polyethylene) enclosing a water reservoir of one gallon minimum capacity indicated by a fill line, an intake air filter, and a self-lubricating motor. The humidifier shall operate on 115 volts, 60 Hz, AC, and shall be capable of continuous operation for at least nine hours without refilling. The humidifier shall emit cool vapor utilizing a minimum of 7 ounces of water per hour.

The humidifier motor shall create a steady flow of cool water vapor through high speed centrifugal action. The motor shall be fuse protected. The electric power cord shall be a 3-conductor type including a ground lead, and terminating in a 3-prong (1 ground) molded male plug that complies with American National Standards Institute (ANSI) Standard C73.11. The power cord shall be strain relieved by a knot in the end which shall withstand a 35 pound pull applied to the power cord with the cord disconnected within the humidifier.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The intake air filter shall be 1/4 inch thick urethane polyester foam secured in an enclosure over the ventilating openings in the vaporizer housing. As an alternate, a suitable filter cartridge, designed to remove dust and pollen, shall be acceptable. Bolts, screws, nuts, and washers shall be corrosion-resisting steel, or chromium-plated brass. The filter shall be easily replaced when it is deemed necessary. Three replacement filters shall be supplied with the humidifier.

The humidifier shall have an Underwriters Laboratories (UL) label or UL listing. A laboratory report from a qualified independent laboratory is acceptable, if the laboratory is adequately equipped and competent to determine compliance with UL standards. The humidifier shall meet the requirements of UL 544, Standards for Safety, Medical and Dental Equipment. The humidifier shall be marked with the electrical rating: voltage, current, power, and cycles.

The exterior surfaces of the humidifier shall be free from burrs or sharp edges. The unit shall be free from leaks, or cracks in reservoir. The humidifier shall be free of grease, oil, or foreign matter.

Workmanship. The portable humidifier shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One humidifier, together with instructions for maintenance and use, as specified, constitutes one unit. Each unit shall be packaged in a suitable commercial container which shall provide adequate protection during shipment and handling.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6530-00-405-5908

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6530-2285

Location: HUMID13C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54981	950120

3 DOCUMENT TITLE
HUMIDIFIER, PORTABLE (Therapeutic)

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MAP	b TELEPHONE (Include Area Code) (1) Commercial (215)737-8431	(2) AUTOVON 444-8431
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041 3466 Telephone (703) 756 2340 AUTOVON 289 2340	

DD Form 1426, OCT 89 Previous editions are obsolete 191

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.