METRIC

A-A-54994
16 December 1994

COMMERCIAL ITEM DESCRIPTION

PHACOEMULSIFICATION VITRECTOMY SYSTEM

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a Phacoemulsification Vitrectomy System that combines anterior and posterior procedures with irrigation, aspiration, vitrectomy and illumination.

Salient Characteristics:

The phacoemulsification vitrectomy system shall be a combined anterior/posterior system with irrigation, aspiration, vitrectomy, illumination, phacoemulsification, phacofragmentation, and air/fluid exchange capabilities, with automated microscissors, two phaco and two frag handpieces, and a bipolar coagulator with cord and footpedal.

The system shall be used with dry compressed air or medical grade nitrogen only, 80 to 100 psi (558.8 to 689.5 kPa), 3.4 scfm (96,288 sccm).

The system shall operate in a temperature range of 50 to 85°F (10 to 30°C), at a relative humidity of 90 percent non-condensing.

The system shall have a cassette capacity of 225 cc (13.73 ci).

The system shall have an unpacked weight of 39.92 kg (88 lbs), height 38.74 cm (15.25 in.), width 47.625 cm (18.75 in.), and depth 43.5 cm (17.125 in.).

The system shall have an accessory connection which is normally closed. If programmed the toggles close when the right top footswitch is pressed a second time. The maximum rating shall be 24 VAC, 1 amp.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The bipolar coagulator shall have a nominal power rating of 7.5 watts at 100 ohms. It shall be adjustable from 2 percent to 100 percent in increments of 2 percent. The duty cycle shall be 75 percent.

The system phacoemulsification/fragmentation power shall be 35 watts maximum at approximately 5k ohms. The frequency shall be 29 kHz, ±2 kHz. Power shall be adjustable from 5 percent to 100 percent in increments of 5 percent. Pulse mode duty cycle shall be 50 percent.

The system shall have a controlled anterior capsulotomy capability of 125 cuts per second with a complex cutting action.

The system shall have vitrectomy capability of 30 to 750 cuts per minute, adjustable in increments of 30 cuts per minute.

The system shall have an anterior aspiration capability of 0 to 550 mmHg, console adjustable from 5 to 550 mmHg in increments of 5 mmHg. The posterior aspiration capability shall be 0 to 400 mmHg, console adjustable from 5 to 400 mmHg in increments of 5 mmHg.

The automatic microscissors shall be capable of 30 to 300 cuts per minute, adjustable in increments of 10 cuts per minute.

The system shall have a fiber-optic illumination bulb of 15 volts, 150 watts.

The electrical characteristics of the system shall be as follows:

110 ±10 percent, 50 ±2 Hz
110 ±10 percent, 60 ±2 Hz
230 ±10 percent, 50 ±2 Hz
230 ±10 percent, 60 ±2 Hz

Workmanship. The phacoemulsification vitrectomy system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One phacoemulsification vitrectomy system, as specified, constitutes one unit. Shall be supplied in a suitable unit carton.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an Acceptable Quality Level (AQL) of 1.0 (percent defective) and an inspection level of S-2. The unit of product for examination purposes shall be one phacoemulsification vitrectomy system.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one phacoemulsification vitrectomy system.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Test shall include, but shall not be limited to, the compliance with specifications.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-371-7062	Phacoemulsification Vitrectomy System

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6515-5743

Location: PHACM17C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54994	941216

3 DOCUMENT TITLE

PHACOEMULSIFICATION VITRECTOMY SYSTEM

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code)
DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MAP	(1) Commercial (215)737-8431 (2) AUTOVON 444-8431
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church, VA 22041 3466 Telephone (703) 756-2340 AUTOVON 289 2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.