INCH-POUND

A-A-54995
16 December 1994

COMMERCIAL ITEM DESCRIPTION

REFRIGERATOR, SOLID STATE BIOLOGICALS

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a portable solid state biologicals refrigerator.

Salient characteristics:

The refrigerator shall be capable of maintaining operating temperatures between 33.8 to 42.8°F (1 to 6°C), inclusive within all areas of the storage compartment. The refrigerator shall be capable of cooling the minimum payload of whole blood units described below from 50°F to 43°F, at an ambient temperature of 81°F, within seven hours. The refrigerator shall be capable of maintaining the payload temperature below 10°C following power loss at 77°F (25°C) ambient temperature for a minimum of six hours providing the appropriate 33.8 to 42.8°F (1 to 6°C) storage temperature existed prior to power failure. It shall be suited for use in all rugged field environments within temperatures ranging from -40°F through 120°F, inclusive.

The unit shall pull down to 50°F (10°C) within 7 hours at an ambient temperature of 77°F (25°C) [after being precooled to no lower than 33.8°F (1°C)], with a full load of blood initially at 77°F (25°C). Internal volume shall be $1Ft^3$ minimum with a minimum payload of 30 whole blood units of 450ml each or 50 packed red blood cell units of 270ml each.

AMSC N/A FSC 4110

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The refrigerator with all accessories in place and containing a minimum payload of 30 whole blood units of 450ml each or 50 packed red blood cell units of 270ml each shall not exceed 125 pounds in weight and 8.5 cubic feet in external volume. Outside dimensions of unit shall not exceed 22.0 inches in depth or 32.0 inches in width (length). Volumes and dimensions do not include recording thermometer. Legs are not required on the refrigerator.

The unit's internal storage capacity shall be capable of storing the minimum payload as described above. It shall have an easily removable basket(s) to facilitate loading and unloading.

The refrigerator shall be equipped with a temperature indicator, and an audible and visual alarm system which will alert the operator that an abnormal temperature condition exists or existed. The temperature indicator and alarm system shall be flush mounted as an integral part of the unit. The alarm shall be activated when the temperature falls outside the acceptable 33.8 to 42.8^{o}F (1 to 6°C) temperature range or in the event of electrical power loss. The alarm shall have a minimum dB rating of 60. The power source for the alarm system shall be separate from that of the refrigerator and shall incorporate a continually recharging battery power source or utilize a standard commercially available battery in the Federal Supply System. The capability to measure battery voltage or indicate low battery voltage shall be included in the alarm system.

The unit shall be certified by an independent testing laboratory to be in compliance with UL 471, Section 49, Rain Test; Section 37, Leakage Current Test; and Section 44, Dielectric Voltage Withstand Test.

The unit shall be supplied with a recording thermometer to provide a continuous hard copy recording of at least one week of storage compartment temperatures under non-transit conditions. The recording thermometer shall record all temperatures between 14 to 68°F (-10 to 20°C), in increments of 2°C as a minimum. The recording thermometer shall be supplied as an integral component of the thermometer alarm system or as a separate accessory item (packaged with the refrigerator). The maximum interval between points of record shall be 5 minutes. If the temperature recorder is supplied as an accessory, it shall have a carrying handle.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The unit shall be capable of operating from 110/220 VAC, 50/60 Hz and 12/24/28 VDC power supplies. AC and DC input power circuits shall be protected via circuit breakers. Control and power circuits shall be isolated from input line power circuit.

A minimum of 2 carrying handles shall be affixed to the refrigerator to facilitate loading and unloading the unit from vehicles and carrying the unit to the operational site.

The refrigerator shall be durably constructed to withstand normal field handling and vibration peculiar to wheeled vehicle movement over roads, trails, and cross-country.

The unit shall be nonreflective "olive drab" in color. The units are not subject to biological and chemical warfare cleaning.

Workmanship. The solid state biologicals refrigerator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One refrigerator, as specified, constitutes one unit. Refrigerator shall be supplied in a suitable unit carton.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be performed on a product demonstration model or first article sample, as follows, to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

1. Examine the unit for removable storage racks. Load the unit with thirty bags of simulated blood of 450ml each. Examine the storage racks for bending or breaking.

2. Weigh the unit loaded with thirty bags of simulated blood of 450ml each. The loaded unit shall weigh no more than 125 pounds.

3. Measure the external volume of the unit. The unit shall have an external volume of 8.5 cubic feet.

4. Examine the unit for power requirements. The unit shall operate on all of the following power sources:

120 VAC, 50 Hz	220 VAC, 60 Hz
120 VAC, 60 Hz	12 VDC
220 VAC, 50 Hz	28 VDC

5. Load the unit with 30 blood bags of 450ml each filled with fluid. Precondition the unit to approximately 39.2^{o}F (4°C). Vary the ambient temperature as follows and record the internal temperature via thermocouple each hour. The unit internal temperature shall stay within the range 33.8 to 42.8^{o}F (1 to 6°C). Power the unit from 120 VAC, 60 Hz.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Time	Ambient	Time	Ambient
0600	90°F (32°C)	1900	114°F (46°C)
0700	91°F (33°C)	2000	108°F (42°C)
0800	95°F (35°C)	2100	105°F (41°C)
0900	101°F (38°C)	2200	102°F (39°C)
1000	106°F (41°C)	2300	100°F (38°C)
1100	110°F (43°C)	2400	98°F (37°C)
1200	112°F (44°C)	0100	95°F (35°C)
1300	116°F (47°C)	0200	94°F (34°C)
1400	118°F (48°C)	0300	93°F (34°C)
1500	119°F (48°C)	0400	92°F (33°C)
1600	120°F (49°C)	0500	91°F (33°C)
1700	119°F (48°C)	0600	90°F (32°C)
1800	118°F (48°C)		

6. Precondition the loaded unit to within the 33.8°F to 42.8°F (1 to 6°C) range. Remove power from the unit. The internal temperature of the unit shall not exceed 10°C for a minimum of 6 hours with an ambient temperature of approximately 77°F (25°C).

7. Load the refrigerator to specified capacity with blood bags filled with blood simulant. Operate the refrigerator until the blood simulant is within the 33.8 to 42.8°F (1°C-6°C) temperature range. Secure the refrigerator to a vibration platform. Excite the refrigerator at a 10 to 50 Hz frequency, with power spectral density of 0.15 g^2/Hz for 20 minutes. Upon completion of excitation, examine refrigerator for satisfactory operation.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. Failure of one unit to pass inspection shall result in failure of the lot. The unit of product for examination purposes shall be one solid state biologicals refrigerator as specified in this description.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Number is covered by this document.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
4110-01-388-9220	Refrigerator, Solid State Biologicals

NOTE 2: Underwriters Laboratory (UL) standards are available from UL, Publications Stock, 333 Pfingsten Road, Northbrook, IL 60062.

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 4110-0491

Location: FRIDG17C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54995	941216

3. DOCUMENT TITLE
REFRIGERATOR, SOLID STATE BIOLOGICALS

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MAP	(1) Commercial (215)737-8431	(2) AUTOVON 444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete. 198

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.