INCH-POUND

A-A-54996
16 December 1994

COMMERCIAL ITEM DESCRIPTION

LIGHT, ULTRAVIOLET, SPECIMEN EXAMINING

(110 Volts, 60 Cycle, AC or Battery Operated)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a portable ultraviolet light, suitable for veterinary use in inspecting food stores for possible contamination or spoilage

Salient characteristics:

This light and carrying case shall be suitable for veterinary use in inspecting food stores for possible contamination or spoilage. The light shall be one integral self contained unit easily held in one hand.

This light shall include, but not be limited to, lamp housing, lamp bulb, reflector, filter, filter frame, ballast, on/off switch, 3 conductor cord with hospital grade plug, provisions for battery operation, and a carrying case.

The lamp housing shall be made of high impact material which will withstand severe knocks. The housing shall withstand 3 drops of 4 feet each onto a concrete floor at any angle without impairing its operating capability or its appearance Bulb and filter may be removed for this test.

The light shall emit long wavelength ultraviolet with peak intensity in the range of 365 nanometers The light intensity shall be at least 200 microwatts/square centimeter when measured 6 inches from the source At 6 inches from the light, the usable viewing area shall be 3 inches by 1-1/2 inches minimum

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The light shall have 95 percent of its visible light rays filtered out.

The lamp bulb shall have a rated life of 6000 hours

If a ballast is used, it shall be vacuum impregnated.

A carrying case shall be provided which shall be suitable for containing the light, all parts and accessories, the two 45 volts batteries and the wiring. This case shall provide complete portability and protection.

The case shall be formed steel, 22-gauge metal (U.S standard gauge), with hinged lid, luggage-type latch, and painted in a green metallic finish.

The case shall allow daylight examination of specimen by providing a lamp holder to position the lamp to shine directly into the chamber; and by providing two hinged access doors, one on one end and the other on an adjacent side. Either of these doors will allow manipulation of a specimen for daylight viewing inside the case under the positioned light

Inside the case an electrical receptacle for plugging the primary cord plug shall be provided to allow operation of the light from two 45 volt batteries through the special battery adapter circuitry provided.

The light shall operate on both 110 volt, 50-60 cycle AC as well as on Direct Current of two 45 volt "B" batteries. A light cord shall be provided, 6 feet to 6-1/2 feet long, terminated with a hospital grade plug. The light shall meet the U L or C S.A electrical requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The nominal overall dimensions of the lamp, exclusive of the primary cord, shall be 11 inches in length, 2 inches in width, and 2-1/2 inches in depth.

The total weight of the ultraviolet lamp shall not exceed 1.35 pounds.

Dimensions of the carrying case shall not exceed 13 inches by 9-1/2 inches by 6 inches

The total weight of the carrying case containing light, all parts and accessories, wiring, not including the batteries, shall not exceed 8.5 pounds.

The light shall be furnished without the batteries but shall be suitable in all respects for use with the batteries. (See Note 2)

Spare Parts In addition to the lamp furnished in the light, one additional lamp shall be furnished as a spare part

Workmanship The light and light system shall be free from defects which detract from its appearance or impair its serviceability

Unit Each (EA) One light, together with all accessories and parts, as specified, constitutes one unit. The carrying case containing the unit shall be packaged in a suitable carton which shall protect contents from damage during shipment and handling.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an Acceptable Quality Level (AQL) of 1 (percent defective) and an inspection level of II The unit of product for examination purposes shall be one light and light system

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one light and light system

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

A drop test shall be performed on both the light alone and the entire light system in the case The lamp housing without bulb and filter shall withstand 3 drops of 4 feet, onto a concrete floor, at any angle, without breaking or impairing its operation.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description The government reserves the right to require proof that the product meets these salient characteristics prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Number is covered by this document.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6530-00-663-2140	Light, Ultraviolet, Specimen Examining (110 Volts, 60 Cycle, AC or Battery Operated)

NOTE 2: The following National Stock Number and Item Identification is referenced by this document.

National Stock Number	Item Identification
6135-00-100-0464	Battery, Nonrechargeable

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6530-2448

Location: LIGHT13C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE	1. DOCUMENT NUMBER: A-A-54996	2. DOCUMENT DATE (YYMMDD): 941216

3. DOCUMENT TITLE: LIGHT, ULTRAVIOLET, SPECIMEN EXAMINING (110 Volts, 60 Cycle, AC or Battery Operated)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME: DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MAP	b. TELEPHONE (Include Area Code) (1) Commercial: (215) 737-8431 (2) AUTOVON: 444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.