MIL-C-36435B
13 November 1975
SUPERSEDING
MIL-C-36435a
14 January 1966

MILITARY SPECIFICATION

CUPS, SPECIMEN

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE.

1.1 Scope. This specification covers lightweight, waterproof, leakproof paper containers, with covers, suitable for use as a container for laboratory specimens.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-636	Boxes, Shipping, Fiberboard.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:35Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material. The specimen cup shall be fabricated of kraft containerboard or bleached kraft plastic coated on the inside.

3.1.1 Coating. The plastic coating shall be a non-toxic thermo plastic (polyethylene) or a copolymer of vinyl chloride, vinyl acetate. The coating shall be free of any disagreeable odor and shall not contaminate the intended end use contents.

3.1.2 Ink. The ink used for markings shall be standard commercial waterproof ink generally used by the food packaging industry.

3.2 Construction. The specimen cup shall be of round, nested construction with sturdy, smooth, rolled brims and recessed bottoms. The cup capacity shall be not less than 8 fluid ounces. The plastic coating shall be on the inside only and shall extend over the rolled rim. The coating shall be a continuous film and show no evidence of staining. The cups shall have sufficient rigidity to prevent collapsing under normal filling and handling conditions.

3.2.1 Body. The cup body shall have the following dimensions:

Outside roll diameter	3-5/8 inches ± 1/16 inches
Covered outside height	2-17/32 inches ± 1/16 inches
Cup depth	1-27/32 inches ± 1/16 inches

The specimen cup shall be a container for laboratory specimens, and shall have a printed legend for contents and date.

3.2.2 Closures. The specimen cup closures shall be the coverall type forming a spillproof seal when seated on the cup. The closures shall have the specified continuous plastic coating on the product contact side.

3.3 Capacity. The specimen cup capacity shall be 8 ounces when tested as specified in 4.4.2.

3.4 Leakage. The specimen cup shall show no signs of leakage when tested as specified in 4.4.1.

3.5 Marking. All markings shall be with ink specified in 3.1.2 All letters shall be not less than 1/8 inch in height.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:35Z
Check the source to verify that this is the current version before use.

3.5.1 Identification marking. The manufacturers name or registered trademark shall be legibly printed on the cup.

3.5.2 Other marking. Other marking shall consist of at least "CONTENTS" and "DATE" with sufficient underlined space for the required specimen information. Additional information printed on the outside of the cup shall be acceptable.

3.6 Workmanship. Workmanship shall be first class throughout. Cups shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, and auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or material may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:35Z
Check the source to verify that this is the current version before use.

4.2 Sampling. The unit of product for inspection purposes shall be the individual cup.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I.

Table I. Sampling for Examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	S-4	2.5
Total (major and minor defects combined)	S-4	6.5
For dimensional examination	S-2	4.0

4.2.2 For tests. Sampling for leakage test shall be conducted in accordance with MIL-STD-105, inspection level S-2. For capacity testing, not less than five samples shall be taken regardless of lot size. Acceptance number shall be zero for all sample sizes.

4.3 Examination. The cups shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in Table II. Examination shall not be restricted to the classified possible defects listed below.

Table II. Classification of Defects

Categories and defects	
Major	
101	Cup not complete with all required component parts.
102	Printing not as specified.
103	Any lid does not seat tightly on body to provide for spillproof seal.
Minor	
201	Identification marking not present or not complete with specified information.

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T08:35Z
Check the source to verify that this is the currentversion before use.

4.4 Tests. Tests shall be conducted as follows and, as necessary, to determine compliance with specification requirements.

4.4.1 Leakage test. Cups shall be placed in their normal position (not inverted) and without closures, filled with cold tap water and allowed to stand for 24 hours. At the end of the test period, the cups shall be visually examined for any signs of leaking.

4.4.2 Capacity test. The cup shall be filled with 8 ounces of water. After filling it shall be possible to seat the closure securely in place.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table III. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table III. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A, B or C as specified (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:35Z
Check the source to verify that this is the current version before use.

5.1.1 Unit of issue. One package containing five hundred (500) cups and covers, as specified, constitutes one unit of issue.

5.1.1.1 Unit package.

5.1.1.1.1 Level A. One unit of issue (500 cups and covers) shall be packaged in a box of appropriate size and constructed in accordance with PPP-B-636, class weather-resistant, grade V3c. Box shall include partitions of suitable design to prevent shifting of, or damage to the contents. As an alternate to the partitions, the cups and covers may be wrapped or bagged in kraft paper or plastic material. Closure, waterproofing and strapping shall be as specified in the appendix of the box specification.

5.1.1.1.2 Level B. One unit of issue (500 cups and covers) shall be packaged as specified in 5.1.1.1.1, except the box shall conform to PPP-B-636, type CF, class domestic. Waterproofing and strapping shall not be required.

5.1.1.1.3 Level C. The subject commodity shall be packaged in substantial commercial containers (export when applicable) of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2 Packing. No further packing required.

5.3 Marking. Each unit package shall be marked as required for the exterior container in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. Intended for use as a container for laboratory specimens.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging (see 5.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:35Z
Check the source to verify that this is the current version before use.

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-027-6155	CUP, SPECIMEN, Paper, Disposable, 3-1/2 inch Diameter, 2-1/2 inch Deep, 8 oz, 500s

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

6.5 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1340
CS

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:35Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE: MIL-C-36435B CUPS, SPECIMEN

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

GPO 1974—603 680 8128 2

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:35Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FO)

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:35Z
Check the source to verify that this is the current version before use.