|INCH-POUND|

MIL-C-36465A
25 June 1989
SUPERSEDING
MIL-C-36465
15 January 1966

MILITARY SPECIFICATION

COT, FOLDING, HOSPITAL

Aluminum

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of lightweight, adjustable, compactly folding, aluminum cot with a canvas cover.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1. Specifications, standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation (see 6.2).

|Beneficial Comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.|

AMSC/NA FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:38Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

V-T-285	Thread, Polyester
QQ-A-200/8	Aluminum Alloy 6061, Bar, Rod, Shapes, Tube and, Wire, Extruded
QQ-A-591	Aluminum Alloy Die Castings
QQ-A-601	Aluminum Alloy Sand Castings
TT-E-527	Enamel, Alkyd, Lusterless
TT-P-1757	Primer Coating, Zinc Chromate, Low-Moisture-Sensitivity
CCC-D-950	Dyeing and Aftertreating Processes for Cotton Cloths
CCC-T-191	Textile Test Methods
PPP-B-601	Boxes, Wood, Cleated-Plywood
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner
PPP-B-636	Boxes, Shipping, Fiberboard

Military

MIL-C-496	Clips, End, Strap
MIL-W-530	Webbing, Textile, Cotton, General Purpose, Natural or in Colors
MIL-B-543	Buckles, Tongueless and Web Strap
MIL-C-5541	Chemical Conversion Coatings on Aluminum and Aluminum Alloys
MIL-C-7219	Cloth, Duck, Nylon
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vapor-proof or Waterproof, Flexible
MIL-T-43566	Tape, Textile, Cotton or Polyester, General Purpose, Natural, or in Colors

STANDARDS

Federal

FED-STD-151	Metals; Test Methods
FED-STD-595	Colors
FED-STD-751	Stitches, Seams and Stitchings

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MIL-STD-794	Parts and Equipment, Procedures for Packaging of

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:38Z
Check the source to verify that this is the current version before use.

(Unless otherwise indicated copies of federal and military specifications, standards, and handbooks are available from the Naval Publications and Forms Center, (ATTN: NPODS), 5801 Tabor Avenue, Philadelphia, PA 19120-5099.)

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

DRAWING

Defense Personnel Support Center (DPSC)

2109 Cot, Folding, Aluminum

(Copies of the drawing may be obtained from: The Defense Personnel Support Center, 2800 So. 20th Street, Philadelphia, PA, 19101, Attn: DPSC-RST).

2.2 Non-Government publications. The following documents form a part of this document to the extent specified herein. Unless otherwise specified, the issues of the documents which are DOD adopted are those listed in the issue of the DODISS cited in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS are the issues of the documents cited in the solicitation (see 6.2).

COLOR ASSOCIATION OF THE UNITED STATES, INC.

United States Army Standard Color Card for the Offical Standardized Shades of Sewing Threads

(Application for copies should be addressed to the: Color Association of the United States, Inc., 200 Madison Avenue, New York 16, N.Y.).

(Non-Government standards and other publications are normally available from the organizations that prepare or distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets, or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:38Z
Check the source to verify that this is the current version before use.

3.1 Materials. Except as otherwise indicated, materials shall conform to the applicable portions of the documents referenced herein. When materials to be used are not specifically designated, they shall be suitable for the purpose intended.

3.1.1 Nylon duck. Nylon duck shall conform to MIL-C-7219, 7.25 oz., class 1, Type III, O.G. 106.

3.1.2 Webbing. Webbing for the tie straps shall conform to type IIb, class 4 of MIL-W-530. Reinforcing webbing shall conform to type II (hard texture), class 4 of MIL-W-530. Webbing shall be vat dyed olive drab No. 7. The webbing shall be made mildew resistant by evenly depositing, within the webbing, a maximum of 1.28 percent copper-8-quinolinolate in lieu of the maximum specified in MIL-W-530.

3.1.3 Polyester thread. Thread for all stitching shall conform to V-T-285, for type II, class 1, letter size FF. The thread shall be vat dyed olive drab shade S-1, conforming to the United States Army Standard Color Card. The finished thread shall show good colorfastness to laundering, chlorine bleaching, and weathering.

3.1.4 Metal components.

3.1.4.1 Endpieces, and legs, and sidepieces. Aluminum tubing, both round and square, shall be aluminum alloy 6062-T6. As an alternate, endpieces and legs shall be aluminum alloy tubing temper T6, conforming with QQ-A-200/8, 1-1/8 inches outside diameter and 0.065 inch wall. Side pieces shall be round cornered, square aluminum alloy tubing temper T6, conforming with QQ-A-200/8, 1-1/8 inches square and 0.065 inch wall.

3.1.4.2 Hinge assembly. Hinge Assembly shall be a heat treated aluminum alloy sand casting, alloy 356 T6 conforming to QQ-A-601, class 3, or as an alternate an aluminum alloy die casting conforming to QQ-A-591, class A 380. Assembly shall be with 3/16 inch by 1-5/16 inch roundhead, monel metal or corrosion-resisting steel, annealed, bright-finished rivets.

3.1.4.3 Leg angle braces. Leg angle braces shall be spring steel wire, 1/4 inch diameter.

3.1.4.4 Rivets. Rivets shall be aluminum alloy 2017T4.

3.1.4.5 Buckles. Buckles shall conform to MIl-B-543 for type II, style 4, class 1, 5/8 inch.

3.1.4.6 End clips. End clips shall conform to MIL-C-496 for type I, 5/8 inch, brass.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:38Z
Check the source to verify that this is the current version before use.

3.1.5 Binding. Edge binding shall be lightweight cotton tape conforming to type I, class 4 of MIL-T-43566. Tape shall be vat dyed olive drab no. 7.

3.2 Construction. Construction shall conform to the requirements specified herein and to drawing 2109. Each unit shall be completely assembled.

3.2.1 Stitches, seams, and stitchings. All seams and stitchings shall be straight and shall conform to FED-STD-751. Stitches shall be type 301, formed by machine, with 8 to 12 evenly spaced stitches per inch. The ends of each continuous row of stitches (including ends due to thread breaks) shall be anchored by backstitches for a length of not less than 1 inch, except where anchorage is provided by another seam or stitching. Thread tension shall be properly maintained so that the lock will be embedded in the center of the material sewed.

3.2.2 Construction Details. Construction shall be in accordance with DPSC drawing 2109, and as hereinafter specified. The removable cover shall be made of one piece nylon duck as shown on the drawing. The overall dimensions of the cot shall be 75-1/2 inches long, 27 inches wide, and 16-3/8 inches high in the set-up flat position. The cot shall measure 39 inches by 27 inches by 4-1/2 inches when folded. The cot shall have two adjustable sections, one on each end, each measuring 18 inches long and adjustable from the horizontal to 45 degrees above the horizontal in steps of 15 degrees. Tolerances shall be as shown on the drawing. The end clips for the tie straps shall be firmly attached without cutting the webbing which shall be inserted full depth and be flat at the point of entry. Turnunders and seams shall be neatly made. No loose or raveled edges shall be exposed. All cutouts shall be reinforced by webbing as shown on the drawing except that the edge binding specified in 3.1.5 shall be an acceptable alternate. In addition the turn back on the sides of the cover shall be more than 3 inches, if necessary, in order to utilize excess material.

3.3 Finish.

3.3.1 Degreasing. All surfaces shall be thoroughly cleaned and chemically degreased prior to chemical treatment.

3.3.2 Aluminum parts. All aluminum parts shall be finished with a chemical film conforming to MIL-C-5541.

3.3.3 Primer. After degreasing and chemical treatment, a primer coat shall be applied to all metal parts except the hardware and leg angle braces. The primer shall be zinc-chromate conforming to TT-P-1757. The primer coat shall form a grounding bond and base for the finishing enamel.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:38Z
Check the source to verify that this is the current version before use.

3.3.4 Enamel. The primer coat shall be followed by one coat of lusterless olive drab enamel conforming to TT-E-527. The resultant finish shall be hard, smooth, and highly resistant to abrasion, chipping, and cracking. The color of the final finish shall be lusterless olive drab conforming to color X34087 of FED-STD-595.

3.3.5 Buckles and end clips. Buckles and end clips shall have a black oxidized finish.

3.4 Identification marking. On the underside of the cover, approximately 4 inches from one end and centrally located, there shall be stenciled, using black waterproof ink or waterproof stencil paint, the name or registered trademark of the manufacturer, the contract or purchase order number, and the letters U.S. The size of printing shall be not less than 1/2 inch high.

3.5 Workmanship. Workmanship shall be first class throughout. Complete cots shall be free from defects which detract from their appearance or may impair their serviceability. Stains on the nylon duck, for use in fabricating the end item, shall not be cause for rejection of the end item. Minor distortion of the tubing at the corner bends shall not be cause for rejection, provided the strength of the frame is not impaired and the dimensional tolerances are not exceeded.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified, in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items shall meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:38Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I.

Table I. Sampling for examination.

For visual examination	Inspection level	AQL (defects per hundred units)
Major defects	II	1.0
Total (major and minor combined) defects	II	4.0
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and table II.

Table II. Sampling for Tests.

Component (unit of product)	Characteristic	Requirement	Test procedure
Nylon duck	Yarn ply	3.1.1	4.4.1.1
Untreated	Yarns per inch		
	Weight		
	Breaking strength		
	Nonfibrous material		
	Air permeability		
Thread	Ply	3.1.3	4.4.1.3
	Length per pound		
	Breaking strength		
	Color		
	Elongation		

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:38Z
Check the source to verify that this is the current version before use.

Table II. Sampling for tests. - Continued.

Component (unit of product)	Charateristic	Requirement	Test procedure
Webbing and Tape	Weight	3.1.2 and	4.4.1.2
	Breaking strength	3.1.5	and
	Weave		4.4.1.2.1
	Texture		
	Acidity		
	Water absorption		
	Size content		
	Color		
	Colorfastness		
	Mildew inhibitor		
	content		
	Mildew resistance		
End clips	Material	3.1.4.6	4.4.1.4
	Finish	3.3.5	4.4.1.4
Buckles	Material	3.1.4.5	4.4.1.4
	Finish	3.3.5	4.4.1.4
Enamel	Flexibility	3.3.4	4.4.1.6
	Knife test		

4.2.3 For chemical analysis of metals. Four ounces of each lot of each metal used in the construction of the cot shall be selected for chemical analysis as specified in 4.4.1.5.

4.2.4 Test panels. Sufficient test panels prepared in accordance with the procedure outlined in 3.4 shall be submitted for test outlined in 4.4.1.6.

4.3 Examination. The cots shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with table III.

Table III. Categories and defects.*

Categories	Defects
Major	
101	Cot not complete with all components.
102	Unit not free from burrs, rough surfaces, and sharp edges.
103	Stitching loose or torn affecting servicability.
104	Cot or any component does not function as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:38Z
Check the source to verify that this is the current version before use.

Table III. Categories and defects.* - Continued.

Categories	Defects
Major	
105	Cover not fabricated from one piece of duck.
106	Cover not constructed so that warp runs parallel to the long axis of the litter.
107	Unselvage edge not hemmed.
108	Width of the cover not within tolerance specified.
109	Strap not complete with all component parts (buckle and end clip).
110	End clip end of strap does not pass through buckle readily.
111	Strap not free from cuts and tears.
Minor	
201	Cot does not rest level and free of wobble on a flat surface.
202	Unit not clean; not free from dirt, oil, and grease.
203	Finish not as specified or not free from scratches and other defects.
204	Identification marking missing or incomplete.
205	Style, design and construction of cover not as shown on drawing.
206	Turnunders, seams, and stitch lines, not straight or evenly spaced.
207	Cover not free from loose or unraveled edges.
208	Stitches less than eight stitches per inch.
209	Stitches not backstitched at the ends.
210	Cover not permanently and legibly marked with the name or registered trademark of the manufacturer, the letters U.S., and contract number with back waterproof ink or paint.
211	Style, design and construction of the strap not as shown on the drawing.

* Examination shall not be restricted to the classified defects listed above.

4.3.2 Dimensional examination. The cots shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted, to determine compliance with specification requirements.

4.4.1 Material testing.

4.4.1.1 Nylon duck. The nylon duck shall be tested in accordance with the test procedures of MIL-C-7219.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:38Z
Check the source to verify that this is the current version before use.

4.4.1.2 Webbing and tape. The webbing and tape shall be tested in accordance with the test procedures of MIL-W-530 ande MIL-T-43566, respectively, except that the performance of the tests for mildew inhibitor content and mildew resistance shall be conducted under the following conditions:

a. If the material conforms to the requirements for mildew inhibitor content, the mildew resistance tests shall not be performed, and the material shall be considered as complying with mildew resistance requirements.

b. If the mildew inhibitor content exceeds the maximum requirements, the material shall be rejected.

c. If the mildew inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew resistance provided it passes the mildew resistance test.

4.4.1.2.1 Mildew resistance. The finished webbing and tape shall be tested in accordance with method 4762 of CCC-T-191. The leaching process shall be omitted.

4.4.1.2.2 Water absorption. The finished webbing and tape shall be tested for water absorption by method 5500; the size of the specimen and the size of the blotting paper shall conform to method 5502 of CCC-T-191.

4.4.1.3 Thread. The thread shall be tested in accordance with the test procedures of V-T-285.

4.4.1.4 Other components. The end clips and buckles shall be tested in accordance with the test procedures of MIL-C-496 and MIL-B-543 respectively.

4.4.1.5 Chemical analysis. The samples selected in accordance with 4.2.3 shall be analyzed in accordance with method 111.1 or 112.1 of Federal Test Method Standard No. 151. In case of dispute, method 111.1 shall be used for referee analysis.

4.4.1.6 Enamel. Tests on enamel shall be conducted in accordance with the procedures outlined in TT-E-527.

4.5 Packaging inspection. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:38Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Preservative applications. Preservatives shall not be used.

5.1.1.2 Unit packages. Each cot shall be packaged in a solid fiberboard or doublefaced corrugated fiberboard carton, of appropriate size and design constructed in accordance with PPP-B-636, Type CF, class domestic. Each carton shall be adequately secured to prevent accidental opening. Closure shall be as specified in the applicable box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for the cot shall conform to the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.2).

5.2.1 Level A. Four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636 class weather-resistant. Closure and strapping shall be specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. Four unit packages shall be packed in accordance with the requirements of MIL-STD-794 for level C.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:38Z
Check the source to verify that this is the current version before use.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in length, 52 inches in width, and 54 inches in height. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. The cot packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. The cot packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

6.1 Intended use. The cot is intended for use in field medical units.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:38Z
Check the source to verify that this is the currentversion before use.

6.2 Acquisition requirements. Procurement documents should specify the following.

(a) Title, number, and date of this specification.

(b) Issue of DODISS to be cited in the solicitation and if required, the specific issue of individual documents referenced (see 2.1 and 2.2).

(c) National Stock Number (NSN).

(d) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification covers the following item shown in the Federal Supply Catalog.

National Stock Number	Item Identification
6530-00-299-8517	COT, FOLDING, HOSPITAL, Aluminum

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

6.5 Subject term (key word) listing.

Collapsible bed
Lightweight, adjustable, cot
Canvas covered cot

6.6. Changes from previous issue. Marginal notations are not used in this revision to identify modifications, additions, corrections, and deletions from the previous issue due to the extensiveness of the changes.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Civil Agency Coordination Activities:

FDA-MPQAS
USPHS

Preparing activity:

DOD-MD

Agent:

DLA-DM

Project Number: 6530-1773

Location:
MILSP/COT/FOLDING/HOSP/H13-14

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:38Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil – Downloaded: 2022-12-21 20:30Z
Check the source to verify that this is the current version before use.