MIL-F-36458B
16 July 1985
SUPERSEDING
MIL-F-36458A
17 October 1972

MILITARY SPECIFICATION

FORCEPS, RUBBER DAM CLAMP, DENTAL

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a forceps for placing and removing rubber dam clamps on teeth.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. Unless otherwise specified, the following specifications and standards of the issue listed in that issue of the Department of Defense Index of Specifications and Standards (DoDISS) specified in the solicitation form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-601	Boxes, Wood, Cleated-plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Setup.

Military

MIL-P-116	Methods of Preservation.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:37Z
Check the source to verify that this is the currentversion before use.

STANDARDS

Federal

FED-STD-151 Metal, Test Methods

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.
MIL-STD-147 Palletized Unit Loads
MIL-STD-794 Parts and Equipment, Procedures for Packaging of.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24016 Forceps, Rubber Dam Clamp, Dental

(Copies of specifications, standards and drawing required by manufacturers in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. The issues of the documents which are indicated as DoD adopted shall be the issue listed in the current DoDISS and the supplement thereto, if applicable.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

E 18 Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
E 92 Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:37Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 General requirements. The forceps shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

3.2 Drawing. The drawing forming a part of this specification is a specification control drawing. The supplier shall be responsible for preparing his own shop drawing. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper function, fit, assembly or operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

3.3 Material. Materials used to fabricate the components of the forceps shall be as specified in GG-I-526. The body and grip of the forceps shall be fabricated from Class 4, type 420 stainless steel, except, as an alternate for the grip, shall be fabricated from Class 3, type 304 stainless steel. The rivet, yoke, and spring all shall be fabricated from an austenitic stainless steel.

3.4 Design and construction. The dental forceps shall be suitable for placing and removing rubber dam clamps on teeth. The forceps shall consist of body (2 pieces), grip (2 pieces), rivet, yoke and spring, and shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24016 for design, dimensions and construction.

3.5 Hardness. After proper stress relieving, type 420 steel shall have a Rockwell hardness of C48-C52. The springs shall be hardened to a spring temper.

3.6 Clamping force. The clamping force exerted by the operative points shall be 8 $\pm$ 4 ounces in the fully-closed position and 4 $\pm$ 1 pounds in the fully extended position when tested as specified in 4.4.3.

3.7 Corrosion resistance. The forceps shall meet the corrosion resistance test as specified in 4.4.4.

3.8 Finish. Each component shall be mirror or bright or satin finish.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:37Z
Check the source to verify that this is the currentversion before use.

3.9 Identification marking. In addition to contractor marking, each forceps shall be marked with the name or trademark of the manufacturer, if different from that of the contractor. Forceps of foreign manufacture shall also be marked "stainless".

3.10 Workmanship. Forceps shall be free from defects which detract from their appearance or impair their serviceability, proper functioning, and intended use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Record. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the components, may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in Tables I and III (see 4.2.1 and 4.3) and Table II (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in Table I. Unit of product for examination purposes shall be one forceps.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

TABLE I. Examination Criteria.

Type Defect	Inspection Level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	4.0
Minor A	II	2.5
Minor B	S-2	6.5

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table II, except that the acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be as indicated in the table.

TABLE II. Test Criteria.

Characteristic	Requirement	Test Procedure	Unit of Product	Inspection Level
Chemical Analysis	3.3	4.4.1	Not Applicable	(See 4.2.3)
Hardness	3.5	4.4.2	Forceps	S-2
Clamping Force	3.6	4.4.3	Forceps	S-2
Corrosion Resistance	3.7	4.4.4	Forceps	S-2

4.2.3 Sampling of raw materials. A sufficient quantity of each lot of raw material used in the fabrication of the components of the forceps shall be selected for chemical analysis.

4.2.4 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The forceps shall be examined to determine compliance with the requirements of this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in Table III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of Defects.

Categories	Defects
Critical	None defined.
Major A	
101	Design and construction not as specified.
102	Lock defective, binds, excessive play.
103	Spring or rivet does not prevent yoke from slipping off the forceps.
104	Yoke does not hold forceps in desired position.
105	Spring fails to return jaws to closed position.
106	Serrations do not easily and positively engage yoke.
Major B	
151	Dimensions not as specified in view AA of drawing.
Minor A	
251	Dimensions (other than 151) not as specified.

4.4 Tests.

4.4.1 Chemical Analysis. Metal analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151.

4.4.2 Hardness test. Hardness shall be determined in accordance with ASTM Standard E 18 (Rockwell) or E 92 (Diamond Pyramid Hardness).

4.4.3 Clamping test. With the forceps rigidly supported in a suitable fixture, apply a force to the operative point opposing forceps spring action and perpendicular to the forceps centerline. The force required to obtain impending motion in both the fully closed and fully open positions, when measured with a spring scale of suitable sensitivity, shall be as specified in 3.6.

4.4.4 Corrosion resistance. The metal components of the forceps shall pass the applicable test(s) for corrosion resistance, as specified in GG-I-526.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Preservation. • Preservation shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning. Forceps shall be cleaned in accordance with MIL-P-116.

5.1.1.2 Drying. Forceps shall be dried in accordance with MIL-P-116.

5.1.1.3 Preservative applications. Preservatives shall not be used.

5.1.1.4 Unit Package. Each forceps shall be packaged in accordance with MIL-P-116, method IC-1.

5.1.1.5 Intermediate package. Six unit packages shall be packed in a box conforming to PPP-B-566, PPP-B-636, type CF, class domestic or PPP-B-676. Closure shall be adequate to prevent accidental opening.

5.1.2 Level C. The level C preservation for forceps shall conform to MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B, and C as specified (see 6.2).

5.2.1 Level A. Sixteen intermediate packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure, strapping, and weatherproofing shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Sixteen intermediate packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for forceps shall conform to MIL-STD-794 requirements for this level.

5.2.4 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Quantities less than 40 cubic feet need not be unitized. Unitized loads shall be uniform in size and quantities to the greatest extent practicable.

5.2.4.1 Levels A and B. Forceps, packed as specified in 5.2.1 and 5.2.2, shall be unitized on pallets conforming to MIL-STD-147.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:37Z
Check the source to verify that this is the currentversion before use.

5.2.4.2 Level C. Forceps, packed as specified in 5.2.3, shall be unitized on pallets as specified in conformance with MIL-STD-147.

5.3 Marking.

5.3.1 Levels A, B, and C. Each unit pack, intermediate package, exterior container, and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog:

Stock Number	Item Identification
6520-00-531-6000	FORCEPS, RUBBER DAM CLAMP, DENTAL

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

Project No. 6520-1792

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:37Z
Check the source to verify that this is the currenversion before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-LA
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-F-36458B	

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.