MILITARY SPECIFICATION

INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, SPECTACLE FRAME, 4 DRAWER

1. SCOPE

1.1 This specification covers a metal insert of four drawers, suitable for insertion in the standard medical service aluminum field chest (see 6.1 and 6.2).

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

QQ-A-250/11	Aluminum Alloy, 6061, Plate and Sheet.
QQ-A-200/9	Aluminum Alloy Bar, Rod, Shapes, and Tube, Extruded, 6063.
QQ-A-250/8	Aluminum Alloy, 5052, Plate and Sheet
TT-E-527	Enamel, Alkyd, Lustreless.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-C-5541	Chemical Films for Aluminum and Aluminum Alloys.
MIL-P-8585	Primer Coating, Zinc Chromate, Low-Moisture-Sensitivity.
MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

/FSC 6545/

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.

STANDARDS

Air Force - Navy Aeronautical

AN-525	Screw, Washer Head.

Federal

FED. Test Method STD. No. 151	Metals; Test Methods.
FED. STD. No. 595	Colors.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage
MS-16686	Medical Services Symbol.
MS-20257	Hinge, Continous (ASG).
MS-20470	Rivet, Solid-Universal Head, Aluminum and Aluminum Alloy.

DRAWING

Defense Personnel Support Center

24000	Insert Cabinet, Medical Instrument and Supply Set Chest, Spectacle Frame, 4 Drawer (with Bill of Materials).

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawing (with Bill of Materials) listed in section 2 and shall meet the requirements shown therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in 4.4.1. When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and suitable for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.

3.1.1 Primer. Primer shall conform to color Y of MIL-P-8585.

3.1.2 Enamel. Enamel shall conform to TT-E-527 and to color X34087 (see 6.4) of No. 595, except that for the marking specified in 3.5, color 37038 (see 6.5) of No. 595 shall be used.

3.2 Component parts. The metal insert of drawers shall consist of a drawer retaining frame, and four drawers. Each drawer shall have a hinged cover and shall be provided with six removable partitions. The metal insert shall have all components and hardware as specified herein and in the drawing and Bill of Materials.

3.3 Construction. Construction shall conform in all details to the requirements specified herein and as shown in the drawing. Construction shall be sound and durable in all respects. Assembly shall be welded on jigs capable of maintaining required tolerances.

3.3.1 Welding. Welding shall be done by qualified welders using appropriate aluminum welding equipment in a qualified process. Spot welds shall be capable of developing the tensile strength of the parent metal. All exposed welds shall be smoothed sufficiently to blend with the adjacent surfaces after the enamel finish specified in 3.4 shall have been applied. Welds shall be finished in a clean and workmanlike manner. All welds shall be clean, nonporous, sound, and finished smooth to match adjacent surfaces. All welds shall be free from pitmarks and imperfections. All surfaces shall be free from sharp edges. Other welding shall be continuous on all surfaces, using welding rod of composition similar to parent metal, and suitable for the purpose intended.

3.4 Finish. After welding has been completed, but before final assembly and attachment of cardholders to drawer fronts, all accessible aluminum alloy surfaces of the insert cabinet, except the working surfaces of the drawer slides, shall be finished as specified herein. All aluminum alloy surfaces shall be thoroughly cleaned free from dirt, oil, grease, welding flux, and loose scale with solvents, detergents, or other processes which have no deleterious effect on the surfaces.

After thorough cleaning and degreasing, surfaces to be enameled shall be prepared for topcoating by one of the following three methods:

(1) Chemical treatment in accordance with MIL-C-5541.
(2) Anodic treatment in accordance with MIL-A-8625
(3) Priming treatment, using primer as specified in 3.1.1, thinned and applied by spray in accordance with the directions specified in 5.3.1 of MIL-P-8585, and air-dried for not less than 30 minutes before topcoating.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.

Topcoating shall be accomplished with one coat of enamel, as specified in 3.1.2, applied by spray and cured by a heat treatment equivalent to baking at 250° Fahrenheit (F.) for not less than 45 minutes. The finish shall be hard, smooth, and highly resistant to abrasion, chipping, cracking, acid, alkali, and solvents.

Cardholders shall be attached to the drawer fronts before topcoating is applied.

3.4.1 Film thickness of paint. Contractor shall endeavor to control film thickness so that the thickness of a single dried coat of paint is 0.3 to 0.7 mil for primer and 1.0 to 1.5 mil for enamel.

3.5 Identification markings. The insert cabinets shall be permanently and legibly marked with enamel as specified in 3.1.2 as follows:

Suitably located on the front of the top drawer, there shall be a 1-inch diameter Medical Services Symbol as shown for type I in MS-16686.

Centrally located on the upper surface of the top frame plate in the order shown below, there shall be the following information:

In vertical block letters 3/8-inch high, the legend: "Insert Cabinet, Medical Instrument and Supply Set Chest, Spectacle Frame, 4 Drawer."

In vertical block numerals 3/8-inch high, the stock number 6545-782-2622."

In vertical letters 1/4-inch high, the name or registered trademark of the manufacturer.

In vertical letters 1/4-inch high, the contract or purchase order number.

3.6 Workmanship. Workmanship shall be first class throughout. Frame and drawers shall be square and free from twist. Drawers shall operate smoothly and easily. All burs shall be removed and all sharp corners broken. Completed insert cabinets shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (percent defective)	Minimum sample size
Component material testing	S1	*	3

* Acceptance number shall be zero (0) for all sample sizes.

4.2.3 For chemical analysis of metals. Four (4) ounces of each lot of each metal used in the construction of the insert cabinets shall be selected for chemical analysis as specified in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.

4.2.4 Component materials. Sufficient samples of the component materials shall be selected from each lot of material used in the manufacture of the insert cabinets to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample size for component materials covered by the subsidiary specifications is not specified, the sample size for each test shall be as indicated in 4.2.2. No test result shall fall below the minimum or exceed the maximum values specified. The sample unit for testing shall be the quantity necessary to perform all required tests one time each.

4.2.5 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The insert cabinets shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical analysis of the metals. The samples selected in accordance with 4.2.3 shall be analyzed in conformance with methods 111 or 112 of Test Method No. 151. In case of dispute, method 111 shall be used for referee analysis.

5. PREPARATION FOR DELIVERY

5.1 Packaging

5.1.1 Unit of issue. One (1) insert cabinet, as specified, constitutes one unit of issue.

5.1.2 Level A [illegible]

5.1.2.1 Unit preparation. All exterior finished surfaces shall be protected by wrapping with kraft paper and cushioning with excelsior pads or flexible corrugated fiberboard or other suitable materials. Wrapping and cushioning material shall be secured by tying with cotton cord or tape, or by sealing with kraft gummed paper tape. Work shall be accomplished in such a manner as to insure against marring, scratching, or damage to the unit.

5.2 Packing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.

5.2.1 Level A.

5.2.1.1 Exterior container. Each unit, prepared as specified in 5.1.2, shall be overpacked in such a manner as to prevent shifting or damage, in an exterior container designed for type 3 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table IV; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Each unit, prepared as specified in 5.1.2, shall be overpacked in such a manner as to prevent shifting or damage, in an exterior container conforming to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 3 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table II; or PPP-B-636, type CF, class domestic, and Method II closure shall be utilized.

5.3 Marking. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION

ORGANIZATION	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*	DATE

DD FORM 1 OCT 64 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD UNDER

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

DEFENSE SUPPLY AGENCY
EFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20th ST
PHILADELPHIA, PA. 19101

OFFICIAL BUSINESS

DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER

ATTN: DIRECTORATE MEDICAL MATERIEL

2800 SOUTH 20TH STREET

PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:36Z
Check the source to verify that this is the current version before use.