MILITARY SPECIFICATION

JAR, SURGICAL DRESSING

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of nesting aluminum surgical dressing jar and cover.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-A-250/2	Aluminum Alloy 3003, Plate and Sheet.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-151	Metals, Test Methods

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center

5038 Jar, Surgical Dressing, Aluminum, Nesting.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Preproduction sample. Unless otherwise specified (see 6.1), prior to the start of production of the contract quantity, one preproduction sample of the surgical dressing jar shall be assembled, completely finished, and tested by the contractor. The sample shall be submitted, together with the contractor's inspection and test reports, to the contracting officer for examination and testing to determine compliance with the requirements of this specification. Written approval of the materials, construction, and performance of the sample shall be obtained from the contracting officer, or his authorized representative, prior to the manufacture and assembly of the remaining quantity to be manufactured. Further production of the item by the contractor, prior to this approval, shall be at the contractor's risk. Accepted preproduction sample shall become the property of the procuring activity and shall be included in the quantity called for in the contract or order. In the event that the preproduction sample is not approved, the contractor shall submit a new sample for approval, or modify the existent sample so as to remove the deficiencies cited.

3.2 Material. Material shall be aluminum alloy conforming to QQ-A-250/2 for temper O.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

3.3 Construction. Construction, design, and dimensions shall conform to this specification and Defense Personnel Support Center Drawing 5038.

3.3.1 Body of jar. Body of the jar shall be seamlessly formed of a single piece of aluminum sheet, which shall not be less than 0.049 inch thick at any point in the finished jar. The bottom shall be concave. One container shall nest easily within another without jamming and shall seat well when placed on top of its cover.

3.3.2 Cover. Cover shall be formed seamlessly from a single piece of aluminum, which shall not be less than 0.048 inch thick at any point in the finished cover. Cover shall have a strap handle of 0.051 inch No. 16, Brown and Sharpe Gage (B and S Gage), thick aluminum, spotwelded to the cover top and positioned as shown on the drawing.

3.4 Finish. The jar body and cover shall be finished smooth. The jar shall be thoroughly cleaned to remove foreign particles such as dirt, grease, graphite, and metal. As a final finish, jar shall be alkaline cleaned and nondyed anodized in accordance with MIL-A-8625 for type II, water rinsed sealed.

3.5 Identification marking. Jar shall be permanently and legibly marked with the name or registered trademark of the manufacturer, the contract or purchase order number, and the letters U.S. Markings shall be impressed on the bottom outside surface of the jar at even pressure so as not to penetrate the inside surface.

3.6 Workmanship. Workmanship shall be first class throughout. Jars shall be free from defects which detract from their appearance or which may impair their serviceability.

3.6.1 Welding. Spotwelding shall be durable and sound in every detail. Spotwelding shall be free of surface burning, cracks, and porosity, and shall be of sufficient strength to pull the spot out of the parent metal when pulled apart. Spot welds on exposed surfaces shall not show objectionable pits or extrusions; when a surface of the metal is thus defaced in fabrication, it shall be filed or ground smooth to prevent unsightly blemishes on the finished surface.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (Defects per hundred units)
For visual examination		
Major defects	II	1.0
Total (major and minor combined) defects	II	4.0
For dimensional examination	S-3	2.5

4.3 Examination. The jars will be examined to determine compliance with all requirements contained in this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Categories and defects*

Major
101 Body and cover of jar not seamlessly formed.
102 Body and cover of jar not free of cracks, sharp projections, and sharp edges.
103 Jar does not nest without jamming.
104 Base of jar does not nest stably in cover depression.
105 Cover does not fit jar; cannot be removed with one hand.
106 Handle not secured to cover.

Minor
201 Surfaces of jar and cover not smooth.
202 Surfaces of jar and cover not anodized.
203 Jar and cover not free of dents and deformed areas.
204 Jar and cover not free of dirt, grease, graphite, and metal.
205 Jar and cover not free of rough grind marks, trimming marks, pits, and scales.
206 Welds on cover not free of surface burning, cracks, and porosity.
207 Welds on cover not filed or ground smooth.
208 Identification marking incomplete, impermanent, or illegible.
209 Inside of jar not free of penetration of the identification markings.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The jars shall be examined for defects in dimensions. Any dimension not within the tolerances specified shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. Tests for chemical analysis shall be conducted in accordance with method 111.1 or 112.1 of Federal Test Method Standard No. 151.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One jar, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

5.1.2 Packaging quantities. The number of units of issue indiated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not applicable	24 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each jar and cover shall be individually wrapped with tissue and packaged in a carton of appropriate size (with the jar cover inverted) constructed in accordance with PPP-B-566 or PPP-B-676 except that commercial colors shall be acceptable. Closure of carton shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 *Level C.* The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).
(c) Where preproduction sample shall be sent (see 3.1).
(d) When preproduction sample is not required (see 3.1)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - IM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Category N

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

Stock No. 118 0301 007 2000

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity (as indicated on reverse hereof).

SPECIFICATION

ORGANIZATION (Of submitter) | CITY AND STATE

CONTRACT NO. | QUANTITY OF ITEMS PROCURED | DOLLAR AMOUNT
$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

a. GIVE PARAGRAPH NUMBER AND WORDING.

b. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES. ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity) | DATE

DD FORM 1426

DD PPSO 13582

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD UNDER

DEFENSE SUPPLY AGENCY
DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20th ST
PHILADELPHIA, PA. 19101

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

OFFICIAL BUSINESS

DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER

ATTN: DIRECTORATE MEDICAL MATERIEL

2800 SOUTH 20TH STREET

PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.