MILITARY SPECIFICATION

KNIVES, PERIODONTAL, KIRKLAND

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers knives of Kirkland design.

1.2 Classification. Knives shall be of the following styles, as specified (see 6.1).

Style 1	Kirkland, No. K-15
Style 2	Kirkland, No. K-16

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical, General Specifications.
QQ-B-626	Brass, Leaded and Non-Leaded; Rod, Shapes, Forgings, and Flat Products with Finished Edges (Bar, Flat Wire and Strip).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Cornered.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center Drawing 24038.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The knives shall be in accordance with applicable requirements of GG-I-526 and as hereinafter specified.

3.2 Material.

3.2.1 Operative ends. The operative ends shall be Class 4, Type 440A, 440B, or 440C stainless steel.

3.2.2 Handles. The handles shall be Class 3, Type 302, 303, or 304 stainless steel. As an alternate, the handles shall be chrome or nickel plated brass and shall meet the requirements of composition 11, half hard brass in accordance with QQ-B-626, except that the maximum lead content shall be 0.6 percent. The handles shall meet the requirements of the boil test.

3.3 Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C54 to C60.

3.4 Style, design and dimensions. The knives shall be of Kirkland design with thin scalene-shaped broad-blades, which shall have a cutting edge around the entire periphery of the blade. The style, design and dimensions of the knives shall conform to Defense Personnel Support Center (DPSC) Drawing 24038 for the applicable style. The knives shall be suitable for use in making initial incisions in dissections of gingival tissue during periodontal treatment.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:40Z
Check the source to verify that this is the current version before use.

3.5 Construction. The knives shall be of two-piece construction. The handles shall be hollow. The operative ends shall be permanently secured to the handles. The finger grip portion of the handles shall be finely serrated or fine diamond knurled.

3.6 Finish.

3.6.1 Operative ends. The operative ends shall have a mirror or highly polished finish blended to form a sharp cutting edge. The cutting edge of the blade of the operative end shall be formed by the junction of the smoothly contoured inner and outer surfaces of the blade.

3.6.2 Handles. The handles shall have a mirror or satin finish except for the serrated or knurled portion. The serrated or knurled portion of the handles shall be free of barbs and shall provide adequate finger grip during use.

3.7 Corrosion resistance. The salt-spray test specified for brass composition in GG-I-526 shall not apply. The boil test shall be applicable to the entire instrument.

3.8 Identification marking. Style 1 shall be marked with the identification "Kirkland K-15" and style 2, "Kirkland K-16" respectively. Each instrument shall be marked on the front face. All marking shall be on the smooth portion of the handle.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:40Z
Check the source to verify that this is the currentversion before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The Acceptable Quality Levels (AQL's) shall be as indicated below:

Type defect	Inspection level	AQL (Percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor defects	II	4.0

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, except that the acceptance number shall be zero for all sample sizes. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Component	Characteristic	Rqmt.	Test Procedure	Inspection level
Operative end	Hardness	3.3 and GG-I-526	GG-I-526	S-2
Operative end	Copper Sulfate	GG-I-526	GG-I-526	S-2
Entire knife	Boil test	3.2.2, 3.7 and GG-I-526	GG-I-526	S-2
Entire knife	Composition	3.2.1 and 3.2.2	GG-I-526	*

* Sufficient amount from each lot of each metal used in the construction of the knives shall be selected for composition analysis.

4.3 Examination. The knives shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with the following classification of defects.

Categories	Defects*
Critical:	None defined

Major A:

101	Style and design not as specified.
102	Instrument not constructed as specified.
103	Operative end finish not as specified.
104	Operative end not free from cracks or burrs.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:40Z
Check the source to verify that this is the current version before use.

Major B:	
121	Operative end dimensions not as specified.
Minor:	
201	Handle finish not as specified.
202	Instrument not permanently and legibly marked with name or trademark of the contractor.
203	Instrument identification not present or properly located.
204	Handle dimensions not as specified.

* Examination shall not be restricted to the classified possible defects listed above.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One knife, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	288 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Unit package. Each knife shall be packaged in a carton of appropriate size conforming to PPP-B-676, except that commercial colors shall be acceptable. The carton shall be provided with blocks or other suitable device to hold the instrument securely and prevent damage to the blade. Closure of carton shall be adequate to prevent spilling of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:40Z
Check the source to verify that this is the currentversion before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Style of knife (see 1.2).
(c) Selection of applicable levels of packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Category N

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(FOLD ON THIS LINE)

(FOLD ON THIS LINE, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:40Z
Check the source to verify that this is the current version before use.