MIL-L-36485
20 May 1966

MILITARY SPECIFICATION

LOOP, EYE LENS, WILDER

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one size of Wilder loop, suitable for use in eye surgery.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
QQ-C-320	Chromium Plating (Electrodeposited).
QQ-N-290	Nickel Plating (Electrodeposited).
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Medical Supply Center

20318	Loop, Eye Lens, Wilder

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The loop shall be fabricated from materials in accordance with GG-I-526, as follows:

Component	Specification Designation	
	Class	Type
Blade and shank	4	420
Handle	8	260 or 270

3.2 Hardness. After proper stress relieving Rockwell hardness of the blade and shank shall be C-40 minimum.

3.3 Style, design and dimensions. Shall be a loop, with serrated and fenestrated blade, of the Wilder design. The style, design and dimensions of the loop shall be in accordance with Drawing 20318.

3.4 Finish. The finish shall be mirror or mirror and satin conforming to the requirements of GG-I-526.

3.5 Additional requirements. Except as specified herein, the loops shall comply with the applicable requirements and tests of GG-I-526.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

3.6 Workmanship. Workmanship shall be first class throughout. Loops shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105,and as indicated in table II.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table I:

Table I. Sampling for Tests

Component	Characteristic	Req.	Test Procedure	Unit of Product	Insp. level	AQL
Blade and Shank	Hardness	3.2	GG-I-526	1-loop	S-2	*
	Copper Sulfate	GG-I-526	GG-I-526	1-loop	S-2	*
Blade, Shank and Handle	Boil Test	GG-I-526	GG-I-526	1-loop	S-2	*
Handle	Salt Spray	GG-I-526	GG-I-526	1-loop	S-2	*
	Plating	GG-I-526	QQ-N-290	1-loop	S-2	*
			QQ-C-320			

* Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:40Z
Check the source to verify that this is the currentversion before use.

4.2.3 Raw materials. Four ounces of each lot of each metal used in construction of the loop shall be selected for chemical analysis.

4.2.4 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

Table II: Sampling for examination

Type Defect	Inspection Level	AQL Percent Defective
Major	II	1.0
Minor A	II	2.5
Minor B	S-2	4.0

4.3 Examination. The loops shall be examined to determine compliance with with all requirements contained in this specification. Examination shall be conducted as specified in table III. Any loops in the sample containing one or more defects shall be rejected, and if the number of defective loops in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected. Examination is not restricted to the defects listed in table III.

Table III: Classification of defects in accordance with MIL-STD-105.

Categories	Defects
Critical:	None defined

Major:

101 Blade not curved as required.
102 Serrations completely or partially absent.
103 Serrations not well defined and deeply cut.
104 Ends of serrations not well chamfered.
105 Blade not uniform and symmetrical in dimension.
106 Blade edges (inner and outer) and remainder of loop not finished smooth and free from surface roughness and sharpness.
107 Joint of shank and handle not secure.
108 Soldered or brazed joints and adjacent surfaces not free of flux.
109 Shank and handle not uniform and symmetrical in dimension.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:40Z
Check the source to verify that this is the currentversion before use.

110 Loop not free of cracks and pits.
111 Loop not free of corrosion and discoloration.
112 Blade and shank not free of plating.
113 Handle not completly plated.
114 Labels or printed marking on unit container nor correct, complete and legible.

Minor A:

201 Finish not mirror or mirror and satin.
202 Finish not free of coarse grind marks and excessively scratched surface areas.
203 Identification marking not complete, permanent and legible.
204 Labels or markings on unit container not free of defacing marks, scratches and tears.

Minor B:

251 Dimensions not within tolerances specified.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One loop, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	360 units

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:40Z
Check the source to verify that this is the current version before use.

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Carton shall be designed with interior fittings or other suitable device to hold the instrument securely and prevent injury to the blade. Closure of carton shall be adequate to prevent spilling of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:40Z
Check the source to verify that this is the currentversion before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

CATEGORY - N

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:40Z
Check the source to verify that this is the currentversion before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Headquarters, Defense Personnel Support Center
Attn: Directorate of Medical Materiel, Code ATT
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:40Z
Check the source to verify that this is the current version before use.