MIL-T-36476
30 March 1966

MILITARY SPECIFICATION

TRACER, GOTHIC ARCH, DENTAL

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers gothic arch tracers suitable for the determination of the correct centric relation of the mandible to the maxilla.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:39Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center

24035	Tracer, Gothic Arch, Dental, Hight.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The gothic arch tracer shall be fabricated of materials conforming to GG-I-526 as hereinafter indicated. The components shall be of class 3 corrosion resisting steel, except the components listed below shall be of class 9 material:

Component	Class 9 material
Central bearing screw.	Nickel silver.
Slotted guide.	Nickel silver.
Locknut.	Nickel plated brass.
Washer stop.	Brass.
Scriber barrel.	Aluminum.
Locknut.	Aluminum.
Spring.	Phosphor bronze.

3.2 Style, design, and dimensions. Gothic arch tracer shall be an instrument that affords convenient means for effectively verifying or establishing centric jaw relation during prosthetic work. Tracer shall be comprised of four parts: an upper tracer, a lower scriber, and two central bearing plates. The bearing plates shall permit retention of the accepted degree of opening and prevent interference by the occlusal rims while tracings are being recorded. The design and dimensions shall comply with Drawing 24035.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:39Z
Check the source to verify that this is the current version before use.

3.3 Finish. The flat surfaces of the gothic arch tracer shall be satin finish. The associated hardware shall be satin or mirror.

3.4 Resistance to boil test. The class 3 components of the gothic arch tracer shall comply with the boil test requirements of GG-I-526.

3.5 Identification marking. The identification marking shall be in compliance with GG-I-526.

3.6 Workmanship. Workmanship shall be first class throughout. The gothic arch tracers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, and investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. Inspection level shall be S-2. The acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 20260-21T08:39Z
Check the source to verify that this is the current version before use.

4.3 Examination. The gothic arch tracer shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One tracer, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	NOT APPLICABLE	48 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method III, within a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Closure of carton shall be adequate to prevent spilling of conter under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:39Z
Check the source to verify that this is the current version before use.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for ype 2 load, and constructed in accordance with PPP-B-585, style 3, for class use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table II or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof ase liner conforming to MIL-L-10547. Closure and sealing shall conform to pplicable paragraphs of appendix thereto. Case liner shall not be required or fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix f applicable box specification. Closure of each fiberboard box shall be as pecified in the appendix, and waterproofing shall conform with 30.4, of PP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with ppendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a re-eiving activity within the continental limits of the United States for storage nd redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, able I, using type III, class 1 plywood; or shall be designed for type 2 load, nd constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, able I; or PPP-B-636, type CF, class domestic and method I closure shall be tilized.

5.2.3 Level C. The subject commodity shall be packed in substantial ommercial containers of the type, size, and kind commonly used for the urpose, so constructed as to insure acceptance and safe delivery by common r other carriers, at the lowest rate, to point of delivery called for in the ontract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance ith MIL-STD-129. When labels are utilized, waterproofing shall be required nly when applicable carton is fabricated of water-resistant material. Contract r purchase order number and name of contractor shall be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:39Z
Check the source to verify that this is the current version before use.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency- DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Category: N

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:39Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64 WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:39Z
Check the source to verify that this is the current version before use.

FOLD

HEADQUARTERS COMMAND
U. S. ARMY ELECTRONICS [illegible]
FORT MONMOUTH, NEW JERSEY 07703

POSTAGE AND FEES PAID
DEPARTMENT OF THE ARMY

OFFICIAL BUSINESS

Commanding General
U. S. Army Electronics Command
ATTN: AMSEL-PP-
[illegible]
Fort Monmouth, New Jersey 07703

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:39Z
Check the source to verify that this is the current version before use.