MIL-W-36453B
16 September 1976
SUPERSEDING
MIL-W-36453A
17 May 1971

MILITARY SPECIFICATION

WORK BOX, DENTAL LABORATORY

This specification is approved for use by all the Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of work box suitable for holding dental prosthetic cases.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Set-Up.

STANDARDS

FEDERAL

FED-STD-66 - Steel, Chemical Composition and Hardenability.
FED-STD-151 - Metals, Test Methods.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, Pa. 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:37Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

10122 - Work Box, Dental Laboratory

(Copies of specifications, standards, drawings and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Materials.

3.1.1 Work box. The work box shall be fabricated from type 302, 304 or 430 corrosion-resisting steel of FED-STD-66.

3.1.2 Card shield. The card shield shall be fabricated from clear methyl methacrylate.

3.2 Style, design and dimensions. The steel work box shall be suitable for holding dental prosthetic cases. The clear plastic card shield shall be suitable for viewing patient identification cards. The style, design and dimensions of the work box shall conform to Defense Personnel Support Center (DPSC) Drawing No. 10122.

3.3 Drawings. The drawings forming a part of this specification are not production drawings. Drawings establish dimensional limitations and control design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions and tolerances within the limitations prescribed on the drawings to provide a product with proper design, configuration, fit and function. No deviations from the dimensional limitations prescribed on the drawings are permissible without prior approval of the contracting officer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

3.4 Construction. Boxes shall be constructed to permit stacking, without binding or excessive play, when tested as specified in 4.3.2. One plastic card shield shall be furnished with each work box, and shall fit in outer and inner channels of box without binding.

3.5 Welding. The welding shall be durable and sound in every detail. Component members shall be carefully and properly aligned prior to welding into permanent position as a straight, rigid assembly. All welding shall be done with appropriate equipment properly regulated and operated by thoroughly trained and experienced welders highly skilled in the operations of stainless steel welding.

3.6 Finish.

3.6.1 Box. All exposed outside surfaces of the boxes shall have a commercial No. 2B or better finish for work boxes fabricated from sheet stock, and a No. 2 or better finish for work boxes fabricated from strip stock. Interior surfaces shall have a uniform mill finish. Boxes shall be thoroughly cleaned to remove foreign matter, such as dirt, grease, and metal. Exposed spot welds shall be ground smooth.

3.6.2 Card shield. The plastic card shield shall have a smooth finish.

3.7 Identification marking. Each box shall be permanently and legibly marked on the bottom outside with the name or registered trade-mark of the manufacturer.

3.8 Workmanship. Workmanship shall be first class throughout. The work box and card shield shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:37Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the steel and plastic materials, may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component or material, the date of testing, the test method, individual test results, the specification requirements, the lot number, the lot size and the sample size.

4.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties). Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

4.2.1 Examination. Boxes shall be examined for defects, including but not limited to those listed in table I.

TABLE I. Classification of defects.

Categories	Defects
Critical	None defined
Major:	
101	Style, design and construction of work box and card shield not as specified.
102	Exposed outside surface of work box not commercial finish as specified.
103	Work box not free from sharp edges.
104	Box rocks on flat surface.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

TABLE I. Classification of defects. (cont'd)

Categories	Defects
Minor:	
201	Surface of work box not free from rough grind marks, pits, scales, and dents.
202	Work box not thoroughly clean.
203	Identification marking not permanently and legibly marked.
204	Exposed spot welds not smooth.
205	Plastic card shield not clear.
206	Plastic card shield not smooth finish.
207	Dimensions of work box and card shield not within tolerances specified.
208	Card shield does not fit in channels as specified.

The inspection level and acceptable quality level (AQL) shall be as listed in table II. The unit of product for examination shall be one box.

TABLE II. Examination criteria.

Type defect	Inspection Level	AQL (Percent defective)
Major	S-2	1.0
Minor	S-2	2.5

4.2.2 Testing. Boxes shall be tested as specified in table III and 4.3. Inspection level and acceptable quality level (AQL) shall be as listed in table III. The unit of product for testing shall be one box.

TABLE III. Test criteria.

Characteristic	Requirement	Test procedure	Inspection level	AQL
Chemical Analysis	3.1.1	4.3.1	**	*
Stacking	3.4	4.3.2	S-2	1.0

* Acceptance number shall be zero for all sample sizes.
** A sufficient quantity of each lot of steel raw material used in the fabrication of the boxes shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

4.2.3 Packaging inspection. The inspection of the unit, intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.3 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.3.1 Chemical analysis. In the absence of certificates of quality for the steel, chemical analysis shall be conducted in accordance with method 111 or method 112 of Federal Test Method Standard 151.

4.3.2 Stacking. The work boxes shall be stacked. The boxes shall be rejected when they do not nest, when nested boxes bind, and when a nested box has more than 3/16-inch horizontal movement.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.1.1 Unit of issue. One work box, as specified, constitutes one unit of issue.

5.1.2 Level A. Each work box shall be clean, dry and packaged in a box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

5.1.3 Level C. Work boxes shall be clean, dry and individually packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.1).

5.2.1 Level A. Twenty-four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and water-proofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Twenty-four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:37Z
Check the source to verify that this is the currentversion before use.

5.2.3 Level C. The packaged boxes shall be packed in shipping containers that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking. Each unit package and exterior container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6520-00-514-2394	WORK BOX, DENTAL LABORATORY

6.3 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to extensiveness of the changes.

Military Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DSA-DM

Project No. 6520-1425

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-W-36453B WORK BOX, DENTAL LABORATORY

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

GPO 1974-- 603-680 6126 2-1

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:37Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:37Z
Check the source to verify that this is the current version before use.