FED. SUP CLASS
6640

REVIEW ACTIVITIES: ARMY - CE
NAVY - AS
AIR FORCE - 82



NOMINAL DIMENSIONAL DATA					
MORTAR DATA				PESTLE DATA	
PART NUMBER	A DISTANCE ACROSS FLATS mm (± 5%)	B INSIDE DIA mm (± 5%)	C HEIGHT mm (±5%)	PART NUMBER	LENGTH mm (± 5%)
MS36105-1	35	26	18	MS36105-4	38
MS36105-2	75	68	37	MS36105-5	75
MS36105-3	100	86	47	MS36105-6	100

MATERIAL: Mortar and pestle shall be of mullite.

DESIGN: Outer surfaces of the mortar walls shall be octagonal with flat planes forming the surfaces of the octagonal shape.

The flat planes shall be tapered to the base.

Base shall have an area large enough to provide a stable resting surface.

Top edge shall be flat.

Inside surface of the mortar shall be curved.

Pestle shall be designed to afford hand comfort in grinding. The grinding end shall be rounded to facilitate grinding.

Mortar and pestle shall be in accordance with Federal Specification NNN-M-560, Type III.

FINISH: Mortar shall be glazed outside, polished inside. Pestle grinding surface shall be polished, handle shall be glazed.

NOTES:
1. For design feature purposes, this standard takes precedence over procurement documents referenced herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 22 SEPT 1960 REVISED (A) 31 MAY 1967

(A) ENTIRE STANDARD REVISED

P.A.	MU	TITLE	MILITARY STANDARD
Other Cust	MS 82	MORTAR AND PESTLE, LABORATORY mullite	MS36105
PROCUREMENT SPECIFICATION NNN-M-560		SUPERSEDES:	SHEET 1 OF 1

DD FORM 1 SEP 67 672-1 (Coordinated)

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:34Z
Check the source to verify that this is the current version before use.