FED. SUP CLASS
6640



PART NUMBER	DIMENSIONAL DATA mm (in.)		
	LENGTH	WIDTH	THICKNESS
MS36252-1	76 (3)	12.7 (1/2)	2-4 (1/16-1/8)

MATERIAL: Copper strip cut from annealed bus bar for electrical conductors that conform to ASTM Specification B187 for Copper Bus Bars and drawn from electrolytic copper ingots conforming to ASTM Specification B5 for Electrolytic Copper Wire Bars.

DESIGN: Strips shall be polished for use in test for determination of free and corrosive sulfur in petroleum products in accordance with ASTM Method D130 and Federal Test Method Standard No. 791.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 19 June 1961 REVISED (A) 27 November 1967 (B) 17 FEB 1982

USER ACTIVITIES: ARMY - CE, MI

REVIEWER ACTIVITIES: ARMY - ME, MD
NAVY - YD
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

(B) ENTIRE STANDARD REVISED

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		CORROSION TEST STRIP copper	MS36252
Procurement Specifications None	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0614
REVISION DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37
Check the source to verify that this is the current version before use.