A-A-51409
16 April 1986

COMMERCIAL ITEM DESCRIPTION

TUBE, DRAINAGE, SURGICAL, THORACIC

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-T-36589B dated 27 April 1976.

This commercial item description covers a series of sterile, radiopaque, plastic thoracic surgical drainage tubes for use following chest surgery.

Salient characteristics:

Material. Tubes shall be manufactured from transparent polyvinyl chloride plastic.

Style and Design. Tubes shall be supplied, as specified, in the following types and sizes:

Type I	-	Right-angled
Type II	-	Straight
Size 1	-	28 French
Size 2	-	32 French
Size 3	-	36 French

Tubes shall be of the dimensions shown in Table I and shall be 19-22 inches in length. Type I tubes shall have the distal 4.5 - 6.0 inches formed at an angle of 90 degrees relative to the major axis of the tube. The extreme distal end of the tube shall be open with smooth and beveled edges. The diameter of the opening on the distal end shall be a minimum of 65 percent of the inside diameter of the tube.

TABLE I. Tube dimensions (in inches)

Size	Wall Thickness	O.D. of tube ± 0.013	O.D. Enlarged Proximal End
(28 Fr)	.039 - .053	.367	.44 - .50
(32 Fr)	.043 - .060	.419	.47 - .53
(36 Fr)	.043 - .065	.472	.51 - .58

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:55Z
Check the source to verify that this is the current version before use.

The proximal end shall be enlarged to form a funnel-type connector. The extreme proximal end shall be beveled. The beveled portion shall be 1.70 ± 0.30 inches in length. The tube shall have a radiopaque stripe throughout its length. The stripe need not extend onto the enlarged proximal end.

The distal end shall have five to six oval eyes which open into the lumen of the tube. The eyes shall be placed symetrically around the tube within 4.0 inches from the distal tip. Eyes shall be uniform in size and shall have a cross-sectional area equal to 60-70 percent of the cross-sectional area of the lumen. The radiopaque stripe shall be interrupted to show the location of the most proximal eye.

Tubes shall bear the manufacturer's name or registered trademark and the French size on the proximal end. Tubes shall be graduated in 2 cm increments beginning 2 cm from the most proximal eye.

X-Ray Opacity test. The drainage tube shall be placed on an unexposed X-ray plate and covered with an aluminum ladder ranging in thickness from at least 1/8 inch to 3/4 inch, in steps of 1/8 inch. X-ray exposure shall be as listed below; accomplished with automatic processing and using appropriate speed screens in a convenient size cassette:

Milliampere-seconds	(MAS)	10
Peak Kilovolts	(KVP)	65 to 70
Focal length distance	(inches)	30 to 40

The tube shall show a definite shadow which clearly delineates the steps 1/4 inch, 3/8 inch, 1/2 inch and 5/8 inch from each other.

Physical Characteristics. Tubes shall have a tensile strength of 1500 psi (min) and an ultimate elongation of 225 percent (min). Tubes shall be tested for tensile strength and elongation in accordance with ASTM D638. Tubes shall have a hardness of 77 ± 5 A Shore Durometer. Material shall be tested for hardness in accordance with ASTM D2240.

Toxicity. Plastic material shall meet the requirements of the United States Pharmacopiea (USP) for Class IV plastics.

Sterility. Tubes shall be tested for sterility in accordance with the USP.

Workmanship. The item shall be processed in such a manner as to be uniform and free from defects that will affect life, serviceability, or appearance. The interior and exterior of the tube shall be smooth and free from debris, foreign matter, and scratches or nicks.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.

Product Quality Level.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing ten tubes, each tube individually packaged sterile in a peel-open type container capable of maintaining the sterility of the contents unless the package is damaged or opened, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or dimensional investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Tubes shall be examined to determine compliance with specification requirements. Sampling shall be in accordance with MIL-STD-105, sampling level II, AQL 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one drainage tube. Sampling shall be S-2 with an AQL of 2.5, unless specified otherwise herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-140-7999	TUBE, DRAINAGE, SURGICAL, THORACIC, Right Angle, 32 French
6515-00-140-8000	TUBE, DRAINAGE, SURGICAL, THORACIC, Right Angle, 36 French
6515-00-763-7366	TUBE, DRAINAGE, SURGICAL, THORACIC, Straight, 32 French
6515-00-763-7367	TUBE, DRAINAGE, SURGICAL, THORACIC, Straight, 28 French
6515-00-866-9073	TUBE, DRAINAGE, SURGICAL, THORACIC, Straight, 36 French

Source: https://assist.dla.mil -- Downloaded: 20260-26T08:55Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC-05043

Project No. 6515-3337

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.