COMMERCIAL ITEM DESCRIPTION

CANNULA, BRAIN, Haynes, 13 Gage

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one size of brain cannulae of the Haynes design suitable for use in neurosurgical procedures.

Salient characteristics:

Style and design. The instrument shall be of the Haynes design with a 13 gage cannula. The distal tip of the cannula shall be closed and rounded to form a blunt tip. The cannula shall be graduated in one centimeter (cm) increments from the distal tip to a working length of 10 cm. The cannula shall have three oval openings spaced 120° apart. The openings shall be located 0.355, 0.684, and 1.010 cm ± 0.025 cm respectively from the distal tip. The cross-sectional area of each opening shall be equal to or greater than the cross-sectional area of the cannula lumen. The proximal end of the instrument shall be a permanently attached Luer lock connector. The Luer lock connector shall meet the dimensional and performance requirements of American National Standards Institute/Health Industry Manufacturer's Association (ANSI/HIMA) Standard MD70.1.

Material. The instrument shall be manufactured from Class 3 austenitic corrosion resisting steel as specified in American Society for Testing and Materials (ASTM) Standard F899.

Passivation. Instruments shall be properly passivated.

Finish. The instrument shall have a mirror or satin finish.

Boil test. The instrument shall show no signs of corrosion when subjected to the following boil test. The instrument shall be thoroughly scrubbed with soap and water, then rinsed with distilled water and dried. The instrument shall then be boiled for 30 minutes in a glass beaker of distilled water. After boiling, the instrument shall remain submerged for another 24 hours. At the end of 24 hours, the instrument shall be inspected for signs of corrosion.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Workmanship. The lumen and exterior shall be free from any chips, slivers, or foreign matter. The finish shall be uniform. There shall be no cracks, hooks, burrs, or any defects which detract from their appearance or impair the serviceability of the instrument.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One brain cannula, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one cannula. Sampling shall be S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-164-8880	CANNULA, BRAIN, Haynes, 13 Gage

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51414

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

52851

Project No. 6515-3531

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.