A-A-51420
18 April 1986

COMMERCIAL ITEM DESCRIPTION

CENTRIFUGE, LABORATORY

The General Services Administration has authorized the use of this commercial item description.

This Commercial Item Description covers one type of table model centrifuge, that is capable of accommodating 1/2 to 100 milliliter (ml) capacity tubes, and suitable for general laboratory use in performing a wide variety of laboratory testing.

Salient characteristics:

Shall be a table model centrifuge, that is operable from 115 volt, 50/60 Hz, AC supply lines and has a maximum power rating of 450 watts.

As a minimum, shall consist of a motor driven 4-place horizontal head, metal housing with cover, power on/off switch, control panel and power cord.

The motor driven horizontal head (unloaded) shall have an adjustable speed range that includes 100 to 2400 RPM. The head shall be designed to accept the following trunnion ring with shield combinations - which shall be commercially available but are not required to be supplied with the centrifuge:

1. 4 each - seven (7) place trunnion ring with 7 shields to accommodate 15 ml. round bottom tubes.

2. 4 each - two (2) place trunnion ring with 2 shields to accommodate 50 ml. tapered bottom tubes.

3. 4 each - one (1) place trunnion ring with shield to accommodate a 100 ml. cone-shaped tube.

4. each - one (1) place trunnion ring with shield to accommodate a 100 ml. pear-shaped tube.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

The design and dimensions of the head shall be such that the following relative centrifugal forces (RCF) can be reached when loaded with the following tube sizes:

Tube size	RCF
15 ml.	1290
50 ml.	1415
100 ml. (cone-shaped)	1305
100 ml. (pear-shaped)	1485

The centrifuge shall include an automatic mode of operation, whereby a desired centrifuging time can be preset, within a range that includes 0 to 99 minutes, and during centrifuging the countdown time to zero shall be indicated on the control panel. In addition, the electrical power to the motor shall be automatically shut off when the countdown reaches "0".

The centrifuge shall include a manual mode of operation, whereby during centrifuging the elapsed time, within a range that includes 0 to 99 minutes, shall be indicated on the control panel. Regardless of the elapsed time, the centrifuge shall continue to operate until manually stopped.

The housing and cover shall have a baked-on enamel finish. A power on/off switch shall be attached to a cutout in the housing. The bottom of the base shall be supplied with rubber or neoprene feet. The cover shall be supplied with a safety device that will not allow the cover to be opened while the centrifuge is in operation.

The control panel shall include the following controls: an automatic/manual mode selection switch, start and stop switches (or push buttons) for operating the centrifuge, timer adjustment automatic mode countdown, variable speed control. The control panel shall include the following visual displays: centrifuge speed, elapsed time during manual mode, and countdown time during automatic mode.

The power line cord shall be attached to a cutout in the housing and the other end shall terminate with a 3-prong male plug.

Workmanship. The centrifuge shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One centrifuge, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspection and testing shall be conducted on each centrifuge to demonstrate its capability for proper operation.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6640-01-119-7870

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-EA
Navy-MS
Air Force-99

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1374

CID-cen-lab-7870

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/45078

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.