A-A-51429
12 May 1986

COMMERCIAL ITEM DESCRIPTION

CALIPER, WAX THICKNESS MEASURING, DENTAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a wax thickness measuring caliper suitable for use in measuring wall thickness of crowns, plastic, baseplates and other semi-rigid dental materials.

Salient characteristics:

The wax thickness measuring caliper shall be suitable for determining the thickness of wax crowns, plastic, baseplates and other semi-rigid dental material. The caliper shall essentially be of five piece construction: operative ends (jaws) with a handle, welded to the end of a metric scale, a sliding handle member with a securely attached knob which moves across the metric scale, and a rivet.

The operative ends (jaws) shall be rounded out, and shall be sufficiently blunt and loosely sprung so as to be suitable for use in measuring semi-rigid dental materials. The measuring beaks shall be configured (like outside calipers) so as to be able to make measurements at any point on a dental wax crown or other semi-rigid materials. The operative ends shall be capable of touching each other at the tips.

The handle members shall have serrations (nominal 35 teeth) along the outside long edges.

The metric scale shall have a range of 0-10 mm capacity legibly and permanently numbered every 0.10 mm.

Overall length (from mid-line of calipers) is 4 $\pm$ 1/16 inches; thickness of jaws shall be 1.50 $\pm$ 0.10 mm; thickness of handle members shall be 1.50 $\pm$ 0.10 mm; thickness of metric scale shall be 1.50 $\pm$ 0.10 mm.

Material. The caliper shall be fabricated of stainless steel or noncorrosive nickel silver.

Workmanship. The caliper shall be free from defects which detracts from its appearance or impair its serviceability.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One caliper, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one caliper. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food,

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from the receiving activity, and shall conform to the applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6520-01-119-6384	CALIPER, WAX THICKNESS MEASURING, DENTAL, 0-10 mm Capacity, 1/10 mm Graduation.

Ordering data. Intermediate/exterior quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Civil Agency Coordinating Activity:
VA-OSS

49126

Project No. 6520-1916

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51429	CALIPER, WAX THICKNESS MEASURING, DENTAT

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5 PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording.

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.