A-A-51431
8 May 1986

COMMERCIAL ITEM DESCRIPTION

BLOWTORCH, ALCOHOL

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-B-36777A.

This commercial item description covers a hand-operated alcohol torch suitable for use in the preparation of various dental prosthesis.

Salient characteristics:

Material Plastic components shall be fabricated from a rigid, thermosetting, black phenolic plastic material suitable for the purpose intended. The wick cap, nozzle, nozzle mounting insert, fixed valve, air reservoir tapped insert, wick holder, filler cap, and front plate, bottom and trigger screws shall be fabricated of brass. Remaining components shall be fabricated of materials suitable in all respects for the purpose intended.

Design and construction. The blowtorch shall consist essentially of a filler cap, wick holder, nozzle holder, alcohol reservoir, air reservoir, and air pump, shown in Figure 1. The blowtorch shall be contoured to fit the operator's hand while the trigger is being operated. Design of the blowtorch shall be such that tightening the thumbscrew valve shall seal its gasket and the accompanying operation of the trigger will produce a constant flow of air and a steadily projected needlepoint flame. A slight turn or opening of the thumbscrew valve on the nozzle holder shall produce an intermittent or brush type flame suitable for spot-heating. The blowtorch shall be rendered suitable for use as an alcohol lamp by turning the nozzle holder out of the flame area. A rubber ring shall be friction-attached to the torch base for shock absorber and protective purposes.

A leather gasket shall be provided on the threaded end of the filler cap to prevent alcohol leakage in the flame burning area. A leather gasket shall be provided on the threaded end of the wick holder to insure a tight seal against alcohol leakage. The wick cap shall rest on the exposed section of the wick, when not in operation, to lessen alcohol evaporation. An extra wick shall be provided to serve as a spare part. The nozzle holder shall consist of a metal male insert partially covered with heat-resisting molded plastic. The alcohol reservoir shall be located in the top section of the torch body and shall be of not less than 90 cubic centimeters (cm^3) capacity. The reservoir shall accommodate denatured alcohol base liquids of the type used to reduce shellac. The air reservoir, having a volume of not less than 105 cm^3, shall be located in the lower part of the torch body. Compressed air shall be forced into the air reservoir by a trigger-operated air pump.

AMSC N/A FSC 6520

DISTRIBUTION A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.




Figure 1. Blowtorch, Alcohol

The blowtorch shown is typical in style and design and is not intended to restrict or preclude the purchase of blowtorches which are otherwise in accordance with the requirements of this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Finish. The plastic material shall be smoothly finished and free of flash. Metal parts shall be smoothly finished and free of burrs.

Workmanship. The blowtorch shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One blowtorch, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product inspection purposes shall be one blowtorch. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

Stock Number	Item Identification
6520-00-561-7150	BLOWTORCH, ALCOHOL

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY

VA-OSS

Project No. 6520-1912

48185

Location: NCR

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51431	BLOWTORCH, ALCOHOL

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording.

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the correct version before use.