A-A-51440
3 September 1986

COMMERCIAL ITEM DESCRIPTION

CARTRIDGE, WATER DEMINERALIZER

The General Services Administration has authorized the use of this commercial item description in preference to military specification MIL-D-37079A.

This Commercial Item Description covers an ion exchange resin cartridge for use with a water demineralizer conforming to National Stock Number 4610-00-976-2420 (Demineralizer, Water, Ion Exchange, 115 Volt, 60 Hz, AC), shall be Crystalab Inc. Model FC-10, or "equal", having the following salient characteristics.

Salient Characteristics:

The ion exchange resin cartridge shall contain a mixture of cation and anion exchange resins. The cation resins shall convert salts (e.g.NaCl) dissolved in water to their corresponding acids. The anion resins shall absorb the acids produced by the cation resins.

Design Construction. The cartridge housing shall be fabricated of plastic and have the shape of a truncated cone. The body of the cartridge shall be filled with ion exchange resins. The top cap and bottom end shall have holes to permit water flow through the ion exchange resins. A porous filter cloth shall be positioned under the cap and over the bottom end. The pore size of the filter cloth shall permit water to flow through the cartridge, and prevent the loss of ion exchange resins from the cartridge. The dimensions of the cartridge shall be such that, when inserted in the shelf opening of a water demineralizer (Crystalab Inc. Model CL-5 or equal), it effects a water-tight seal around the shelf cut-out.

Performance. The ion exchange resin cartridge shall have a minimum absorption capacity of 124 grams of "Sodium Chloride" (NaCl). The water filtration shall be such that the deionization of 24 liters of water, consisting of 3 grams of NaCl in 9 liters of water, results in the initial 1-1/2 liters of water accumulated in the water demineralizer containing less than 1 part per million (ppm) of dissolved salts, and the balance of the water not exceeding 5 ppm of dissolved salts.

Workmanship. The cartridge shall be free from defects which detract from its appearance or impair its serviceability.

AMSC N/A FSC 4610

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the currentversion before use.

Unit. Package (PG). One package containing 6 cartridges, as specified, constitutes one unit.

Inspection/test date. The inspection/test date shall be 36 months from the date of manufacture.

Date of delivery. Not less than 33 months of the inspection/test dating period shall remain at the time of delivery to the Government.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspection and testing. The water demineralizer shall be inspected and tested to determine compliance with specification requirements.

Sampling for inspection and testing. Sampling for inspection and testing shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory Requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping, conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following company to meet the salient characteristics of the item described herein: CRYSTALAB Incorporated Model FC-10.

Note: The following National Stock Number is covered by this document:

4610-00-921-5622

The following NSN, covers the demineralizer water, Ion exchange, that is to be used with this item: 4610-00-976-2420.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.

A-A-51440

MILITARY INTERESTS:

Custodians:

Army-MB
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVILIAN AGENCY COORDINATING ACTIVITY

VA-OSS

Project No. 4610-0109

Location: cartridge-water-dem-5622

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.